

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR POLICE DEPT
DATE 10-2-14 TIME 1714
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

W.T. Ryan

SIGNATURE W.T. Ryan 1100

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
10/02/2014 17:14:31

Standard Lot#: 32012100A1
Last Changed By: W. WILL #974

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 10/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:16:48
Diagnostic Test	Pass	17:17:23
Air Blank	0.000	17:17:50
0.100 Cal Check	0.098	17:18:11
Air Blank	0.000	17:18:41
Subject Test	0.000	17:19:00
Air Blank	0.000	17:19:28
Wait	ABT*	
Air Blank	0.000	17:20:08

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

10/02/2014

17:20:32

Standard Lot#: 32012100A1

Last Changed By: W. WILL

#974

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 10/02/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	17:21:15
Air Blank	RFI**	17:21:38

*Improper Sample

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

10/02/2014

17:22:10

Standard Lot#: 32012100A1

Last Changed By: W. WILL

#974

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 10/02/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	17:22:50
Air Blank	0.000	17:23:19

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

10/02/2014

17:23:55

Standard Lot#: 32012100A1

Last Changed By: W. WILL

#974

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 10/02/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:24:48
Diagnostic Test	Pass	17:25:23
Air Blank	0.000	17:25:51
0.100 Cal Check	0.097	17:26:12
Air Blank	0.000	17:26:41
Subject Test	RFI*	17:28:37
Air Blank	RFI*	17:28:51

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

10/02/2014

17:29:10

Standard Lot#: 32012100A1

Last Changed By: W. WILL

#974

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 10/02/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:30:05
Diagnostic Test	Fail*	17:30:15
Air Blank	RFI**	17:30:22

* Analytical Stability Test

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
10/02/2014 17:30:42

Standard Lot#: 32012100A1
Last Changed By: W. WILL #974

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 10/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:31:36
Diagnostic Test	Pass	17:32:11
Air Blank	0.000	17:32:38
0.100 Cal Check	0.097	17:32:59
Air Blank	0.000	17:33:28
Subject Test	0.024*	17:36:35
Air Blank	0.000	17:37:06
Air Blank	0.000	17:37:34
Subject Test	XXX**	17:38:12
Air Blank	0.000	17:38:48

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

10/02/2014

17:40:07

Standard Lot#: 32012100A1

Last Changed By: W. WILL #974

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	17:40:39
0.100 Cal Check	0.096	17:40:59
Air Blank	0.000	17:41:29

Preventative Maintenance
Performed -- Timer Reset