

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR P.D.
DATE 4-2-14 TIME 1033
INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.10AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 103 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan 1100

DPS Form Exh G-2 (Rev 05-01)

NO PROBLEMS OBSERVED

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

04/02/2014

10:33:43

Standard Lot#: 13712100A2

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 10:34:26
0.100 Cal Check 0.103 10:34:47
Air Blank 0.000 10:35:17

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W. T. RYAN #1100 AGENCY GOODYEAR P.D.

DATE 4-2-14 TIME 1037

INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S. LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 102 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

NO PROBLEMS

SIGNATURE W. T. Ryan 1100

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
04/02/2014 10:37:44

Standard Lot#: 13712100A2
Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 04/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:39:02
Diagnostic Test	Pass	10:39:37
Air Blank	0.000	10:40:04
0.100 Cal Check	0.103	10:40:25
Air Blank	0.000	10:40:55
Subject Test	0.000	10:41:18
Air Blank	0.000	10:41:46
Wait	ABT*	
Air Blank	0.000	10:42:22

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
04/02/2014 10:42:48

Standard Lot#: 13712100A2
Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 04/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:43:38
Air Blank	0.000	10:44:07

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
04/02/2014 10:44:58

Standard Lot#: 13712100A2
Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 04/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:45:47
Air Blank	0.000	10:46:15

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

04/02/2014

10:46:57

Standard Lot#: 13712100A2

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 04/02/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:47:59
Diagnostic Test	Pass	10:48:33
Air Blank	0.000	10:49:01
0.100 Cal Check	0.103	10:49:22
Air Blank	0.000	10:49:52
Subject Test	0.126*	10:52:59
Air Blank	0.000	10:53:31
Air Blank	0.000	10:53:58
Subject Test	XXX**	10:54:22
Air Blank	0.000	10:55:04

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

04/02/2014

10:56:59

Standard Lot#: 13712100A2

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 10:57:30
0.100 Cal Check 0.102 10:57:50
Air Blank 0.000 10:58:21

Preventative Maintenance
Performed -- Timer Reset

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

04/02/2014

10:58:38

Standard Lot#: 13712100A2

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	10:59:09
0.100 Cal Check	0.102	10:59:29
Air Blank	0.000	11:00:00

Preventative Maintenance
Performed -- Timer Reset



ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER MAINTENANCE REPORT

AGENCY Goodyear Police Department
Goodyear, AZ 85338

CR NO. 1800003120002

Page 1 of 1

OFFICER RYAN, #1100

DATE April 24, 2014

INSTRUMENT:

80000312

INSTRUMENT LOCATION:

Goodyear Police Department

MAINTENANCE REQUESTED:

Repair and Preventative Maintenance

MAINTENANCE PERFORMED:

Parts Replaced:

Flow Sensor

Preventative Maintenance Performed:

Inspection, Calibration, Verification

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

JOSEPH M SLOWINSKI, #6101
Electronics Technician
Central Regional Crime Laboratory
P. O. Box 6638
Phoenix, AZ 85005
(602) 223-2394



THE UNIVERSITY OF CHICAGO
LIBRARY



PHYSICS DEPARTMENT
5720 S. UNIVERSITY AVENUE
CHICAGO, ILLINOIS 60637
TEL: 773-936-3700

PHYSICS DEPARTMENT
5720 S. UNIVERSITY AVENUE
CHICAGO, ILLINOIS 60637
TEL: 773-936-3700

PHYSICS DEPARTMENT
5720 S. UNIVERSITY AVENUE
CHICAGO, ILLINOIS 60637
TEL: 773-936-3700

PHYSICS DEPARTMENT
5720 S. UNIVERSITY AVENUE
CHICAGO, ILLINOIS 60637
TEL: 773-936-3700

PHYSICS DEPARTMENT
5720 S. UNIVERSITY AVENUE
CHICAGO, ILLINOIS 60637
TEL: 773-936-3700