

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR POLICE DEPT  
DATE 10-2-14 TIME 1714  
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

W.T. Ryan

SIGNATURE W.T. Ryan 1100

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/02/2014 17:14:31

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 10/02/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:16:48
Diagnostic Test	Pass	17:17:23
Air Blank	0.000	17:17:50
0.100 Cal Check	0.098	17:18:11
Air Blank	0.000	17:18:41
Subject Test	0.000	17:19:00
Air Blank	0.000	17:19:28
Wait	ABT*	
Air Blank	0.000	17:20:08

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/02/2014 17:20:32

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 10/02/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	17:21:15
Air Blank	RFI**	17:21:38

\*Improper Sample  
\*\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/02/2014 17:22:10

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 10/02/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	17:22:50
Air Blank	0.000	17:23:19

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/02/2014 17:23:55

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 10/02/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:24:48
Diagnostic Test	Pass	17:25:23
Air Blank	0.000	17:25:51
0.100 Cal Check	0.097	17:26:12
Air Blank	0.000	17:26:41
Subject Test	RFI*	17:28:37
Air Blank	RFI*	17:28:51

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/02/2014 17:29:10

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 10/02/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:30:05
Diagnostic Test	Fail*	17:30:15
Air Blank	RFI**	17:30:22

\* Analytical Stability Test  
\*\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/02/2014 17:30:42

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 10/02/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:31:36
Diagnostic Test	Pass	17:32:11
Air Blank	0.000	17:32:38
0.100 Cal Check	0.097	17:32:59
Air Blank	0.000	17:33:28
Subject Test	0.024*	17:36:35
Air Blank	0.000	17:37:06
Air Blank	0.000	17:37:34
Subject Test	XXX**	17:38:12
Air Blank	0.000	17:38:48

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

10/02/2014

17:40:07

-----  
Standard Lot#: 32012100A1

Last Changed By: W. WILL #974  
-----

QAS: W. RYAN #1100

GOODYEAR PD  
-----

Operator: W. RYAN #1100

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	17:40:39
0.100 Cal Check	0.096	17:40:59
Air Blank	0.000	17:41:29

Preventative Maintenance  
Performed -- Timer Reset



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. Ryan 1100 AGENCY GOODYEAR P.D.

DATE 10-29-14 TIME 1301 HRS

INTOXILYZER SERIAL # 90-001858 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan 1100

DPS Form Exh G-2 (Rev 05-01)

NO PROBLEMS

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/29/2014 13:01:00

-----  
Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974  
-----

QAS: W. RYAN #1100  
GOODYEAR PD  
-----

Operator: W. RYAN #1100  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	13:01:38
0.100 Cal Check	0.098	13:01:59
Air Blank	0.000	13:02:28

Preventative Maintenance  
Performed -- Timer Reset