

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR POLICE DEPT

DATE 12-24-14 TIME 1213

INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD GOODYEAR, AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS
NO PROBLEMS NOTED

SIGNATURE W.T. Ryan 1100

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
12/24/2014 12:13:02

Standard Lot#: 32012100A1
Last Changed By: W. WILL #974

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 12/24/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:14:30
Diagnostic Test	Pass	12:15:05
Air Blank	0.000	12:15:32
0.100 Cal Check	0.098	12:15:53
Air Blank	0.000	12:16:22
Subject Test	0.000	12:16:52
Air Blank	0.000	12:17:20
Wait	ABT*	
Air Blank	0.000	12:18:11

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
12/24/2014 12:19:36

Standard Lot#: 32012100A1
Last Changed By: W. WILL #974

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 12/24/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	12:20:16
Air Blank	0.000	12:20:45

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
12/24/2014 12:22:19

Standard Lot#: 32012100A1
Last Changed By: W. WILL #974

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 12/24/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	12:22:52
Air Blank	RFI*	12:23:04

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

12/24/2014

12:24:49

Standard Lot#: 32012100A1

Last Changed By: W. WILL

#974

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 12/24/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:25:42
Diagnostic Test	Pass	12:26:17
Air Blank	0.000	12:26:44
0.100 Cal Check	0.098	12:27:05
Air Blank	0.000	12:27:35
Subject Test	0.135*	12:30:42
Air Blank	0.000	12:31:14
Air Blank	0.000	12:31:41
Subject Test	XXX**	12:32:15
Air Blank	0.000	12:32:49

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

12/24/2014

12:35:40

Standard Lot#: 32012100A1

Last Changed By: W. WILL #974

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	12:36:09
0.100 Cal Check	0.098	12:36:30
Air Blank	0.000	12:36:59

Preventative Maintenance
Performed -- Timer Reset