

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR P.D.

DATE 8-20-15 TIME 1442

INTOXILYZER SERIAL # BD-001858 LOCATION 111 S LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check. RESET CLOCK from 1437 TO 1439 HRS  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES W NO    

COMMENTS

---

---

---

---

SIGNATURE W.T. Ryan #1100

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

08/20/2015

14:42:55

---

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 08/20/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	14:44:14
Diagnostic Test	Pass	14:44:49
Air Blank	0.000	14:45:16
0.100 Cal Check	0.098	14:45:37
Air Blank	0.000	14:46:07
Subject Test	0.000	14:46:23
Air Blank	0.000	14:46:50
Wait	ABT*	
Air Blank	0.000	14:47:34

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

08/20/2015

14:47:51

---

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 08/20/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	14:48:29
Air Blank	0.000	14:48:58

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

08/20/2015

14:49:22

---

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 08/20/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	14:49:59
Air Blank	0.000	14:50:27

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

08/20/2015

14:50:55

---

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 08/20/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	14:51:48
Diagnostic Test	Pass	14:52:23
Air Blank	0.000	14:52:50
0.100 Cal Check	0.099	14:53:11
Air Blank	0.000	14:53:40
Subject Test	0.067*	14:56:48
Air Blank	0.000	14:57:19
Air Blank	0.000	14:57:47
Subject Test	XXX**	14:58:27
Air Blank	0.000	14:59:00

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

08/20/2015

14:59:34

-----  
Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      15:00:04  
0.100 Cal Check                0.098                      15:00:25  
Air Blank                      0.000                      15:00:55

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR P.D.

DATE 8-20-15 TIME 1441

INTOXILYZER SERIAL # 80-001858 LOCATION 111 S LITCHFIELD RD GOODYEAR AZ

(S) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- (S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (S) 4. Air blank completed.
- (S) 5. Calibration check completed. Test results 0. 098 AC.
- (S) 6. Air blank completed.
- (S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan #1100

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

08/20/2015

14:41:03

-----  
Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      14:41:36  
0.100 Cal Check                0.098                      14:41:57  
Air Blank                      0.000                      14:42:27

Preventative Maintenance  
Performed -- Timer Reset