

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR P.D.

DATE 9-19-15 TIME 1701

INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD. GOODYEAR AZ

(S) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

(S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(S) 4. Air blank completed.

(S) 5. Calibration check completed. Test results 0. _____ AC.

(S) 6. Air blank completed.

(S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan 1100

DPS Form Exh G-2 (Rev 05-01)

NO ISSUES

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

09/19/2015

17:01:54

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 17:02:25
0.100 Cal Check 0.099 17:02:46
Air Blank 0.000 17:03:16

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUELMAN AGENCY GOOSEVILL PD
DATE 9/27/19 TIME 0413
INTOXILYZER SERIAL # 80-001858 LOCATION 197 N LITCHFIELD RD

- (~~1~~) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (~~2~~) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (~~3~~) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (~~4~~) 4. Air blank completed.
- (~~5~~) 5. Calibration check completed. Test results 0. 099 AC.
- (~~6~~) 6. Air blank completed.
- (~~7~~) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE AG 1144

INTOXILYZER 8000

Location: 190 N LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

09/27/2015

04:13:35

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:14:06
0.100 Cal Check	0.099	04:14:27
Air Blank	0.000	04:14:57

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. GILBERT AGENCY GOOD YEAR PD
DATE 9/26/15 TIME 1517
INTOXILYZER SERIAL # 80-001858 LOCATION 190 N LEECHCREEK RD

- (P) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (P) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (P) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (P) 4. Air blank completed.
- (P) 5. Calibration check completed. Test results 0. 098 AC.
- (P) 6. Air blank completed.
- (P) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE M. G. Gilbert

INTOXILYZER 8000

Location: 190 N LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

09/26/2015

18:17:14

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 18:17:44
0.100 Cal Check 0.098 18:18:05
Air Blank 0.000 18:18:35

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUELBANUE 1144 AGENCY GOODYEAR PD
DATE 9/26/15 TIME 1746
INTOXILYZER SERIAL # PD-001858 LOCATION 1111 S LITCHFIELD RD

- (D) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (A) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(g) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(S) 4. Air blank completed.
(S) 5. Calibration check completed. Test results 0.099 AC.
(S) 6. Air blank completed.
(S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE AK 1144

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

09/26/2015

17:46:47

Standard Lot#: 25913100A2

Last Changed By: W. RYAN #1100

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	17:47:24
0.100 Cal Check	0.099	17:47:45
Air Blank	0.000	17:48:15

Preventative Maintenance
Performed -- Timer Reset