

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR POLICE DEPT
DATE 11-18-15 TIME 2009 HRS
INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES WTR NO _____

COMMENTS

SIGNATURE W.T. Ryan 1100

DPS Form Exh G-4 (Rev 05-01)

NO PROBLEMS NOTED

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:09:19

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/18/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 20:10:26 |
| Diagnostic Test | Pass | 20:11:01 |
| Air Blank | 0.000 | 20:11:28 |
| 0.100 Cal Check | 0.101 | 20:11:49 |
| Air Blank | 0.000 | 20:12:19 |
| Subject Test | 0.000 | 20:12:35 |
| Air Blank | 0.000 | 20:13:02 |
| Wait | ABT* | |
| Air Blank | 0.000 | 20:13:46 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:14:09

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/18/1959.

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 20:14:54 |
| Air Blank | 0.000 | 20:15:22 |

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:15:45

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/18/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 20:16:23 |
| Air Blank | 0.000 | 20:16:52 |

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:17:17

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/18/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 20:18:45 |
| Diagnostic Test | Pass | 20:19:20 |
| Air Blank | 0.000 | 20:19:47 |
| 0.100 Cal Check | 0.101 | 20:20:08 |
| Air Blank | 0.000 | 20:20:38 |
| Subject Test | 0.011 | 20:20:53 |
| Air Blank | 0.000 | 20:21:23 |
| Wait | ABT* | |
| Air Blank | 0.000 | 20:22:39 |

*Sequence Aborted

READ BREATH SAMPLE
READ DEFICIENT SAMPLE

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:22:46

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/18/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 20:23:43 |
| Diagnostic Test | Pass | 20:24:18 |
| Air Blank | 0.000 | 20:24:46 |
| 0.100 Cal Check | 0.101 | 20:25:07 |
| Air Blank | 0.000 | 20:25:37 |
| Subject Test | 0.011* | 20:28:44 |
| Air Blank | 0.000 | 20:29:16 |
| Air Blank | 0.000 | 20:29:43 |
| Subject Test | XXX** | 20:30:07 |
| Air Blank | 0.000 | 20:30:41 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:32:05

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 20:32:36
0.100 Cal Check 0.101 20:32:57
Air Blank 0.000 20:33:27

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR P.D.

DATE 11-18-15 TIME 2006

INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0. 101 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan 1100

DPS Form Exh G-2 (Rev 05-01)

NO PROBLEMS NOTED

INTOXILYZER' 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

11/18/2015

20:06:41

Standard Lot#: 16014100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 20:07:55
0.100 Cal Check 0.101 20:08:15
Air Blank 0.000 20:08:46

Preventative Maintenance
Performed -- Timer Reset