

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. Ryan #1100 AGENCY GOODYEAR Police Dept

DATE 1-11-17 TIME 2353

INTOXILYZER SERIAL # 00-001058 LOCATION 1111 S LITCHFIELD RD, GOODYEAR AZ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan #1100

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/11/2017

23:53:49

Standard Lot#: 28515100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	23:54:19
0.100 Cal Check	0.098	23:54:40
Air Blank	0.000	23:55:10

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR POLICE DEPT

DATE 1-12-17 TIME 0002

INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE W.T. Ryan #1100

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/12/2017

00:02:40

Standard Lot#: 28515100A1

Last Changed By: W. RYAN

#1100

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/12/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:03:35
Diagnostic Test	Pass	00:04:10
Air Blank	0.000	00:04:37
0.100 Cal Check	0.098	00:04:58
Air Blank	RFI*	00:05:15
Air Blank	0.000	00:05:43

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/12/2017

00:06:05

Standard Lot#: 28515100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/12/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:06:56
Diagnostic Test	Pass	00:07:31
Air Blank	0.000	00:07:58
0.100 Cal Check	0.098	00:08:19
Air Blank	0.000	00:08:49
Subject Test	0.000	00:09:08
Air Blank	0.000	00:09:35
Wait	ABT*	
Air Blank	0.000	00:10:13

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/12/2017

00:10:50

Standard Lot#: 28515100A1

Last Changed By: W. RYAN

#1100

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/12/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	00:11:28
Air Blank	0.000	00:11:56

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/12/2017

00:12:14

Standard Lot#: 28515100A1

Last Changed By: W. RYAN

#1100

QAS: W. RYAN

#1100

GOODYEAR PD

Operator: W. RYAN

#1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/12/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	00:12:54
Air Blank	0.000	00:13:22

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/12/2017

00:13:37

Standard Lot#: 28515100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/12/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:14:35
Diagnostic Test	Pass	00:15:10
Air Blank	0.000	00:15:37
0.100 Cal Check	0.098	00:15:58
Air Blank	0.000	00:16:27
Subject Test	0.081*	00:19:35
Air Blank	0.000	00:20:07
Air Blank	0.000	00:20:34
Subject Test	XXX**	00:21:03
Air Blank	0.000	00:21:38

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

01/12/2017

00:22:11

Standard Lot#: 28515100A1

Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	00:22:41
0.100 Cal Check	0.098	00:23:02
Air Blank	0.000	00:23:32

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUEVORGIAN 1144 AGENCY 6000 YERAN PO
DATE 1/17/17 2017 TIME 1404
INTOXILYZER SERIAL # 80-001858 LOCATION 6PO

- (D) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (D) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (D) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (D) 4. Air blank completed.
- (D) 5. Calibration check completed. Test results 0. 099 AC.
- (D) 6. Air blank completed.
- (D) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE AG

INTOXILYZER 8000
Location: 11 S 145TH AVE
Serial Number: 80-001858
Core Version: 8105.48
01/17/2017

14:04:55

Standard Lot#: 28515100A1
Last Changed By: W. RYAN #1100

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:05:26
0.100 Cal Check	0.099	14:05:46
Air Blank	0.000	14:06:16

Preventative Maintenance
Performed -- Timer Reset