

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR POLICE DEPT  
DATE 12-13-17 TIME 2325  
INTOXILYZER SERIAL # 90-000312 LOCATION 11 N 145TH AVE GOODYEAR AZ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

W.T. Ryan #1100

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/13/2017

23:19:33

-----  
Standard Lot#: 14816100A1

Last Changed By: W. RYAN #1100

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	23:21:32
0.100 Cal Check	0.097	23:21:53
Air Blank	0.000	23:22:23
Air Blank	0.000	23:24:52

\*Input Timed Out

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/13/2017

23:25:12

-----  
Standard Lot#: 14816100A1

Last Changed By: W. RYAN #1100

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      23:25:43  
0.100 Cal Check                0.098                      23:26:04  
Air Blank                      0.000                      23:26:34

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W. T. RYAN #1100 AGENCY GOODYEAR POLICE DEPT

DATE 12-13-17 TIME 2348 HRS

INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 095 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

SIGNATURE W. T. Ryan #1100

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/13/2017

23:48:13

---

Standard Lot#: 14816100A1

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 12/13/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	23:49:25
Diagnostic Test	Pass	23:50:00
Air Blank	0.000	23:50:27
0.100 Cal Check	0.097	23:50:48
Air Blank	0.000	23:51:18
Subject Test	0.000	23:51:54
Air Blank	0.000	23:52:21
Wait	ABT*	
Air Blank	0.000	23:53:08

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
12/13/2017 23:55:03

---

Standard Lot#: 14816100A1  
Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: EST,TEST  
DOB: 12/13/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	23:55:54
Air Blank	0.000	23:56:22

\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/13/2017

23:57:08

---

Standard Lot#: 14816100A1

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 12/13/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	23:57:57
Air Blank	RFI*	23:58:05

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/14/2017

00:00:58

---

Standard Lot#: 14816100A1

Last Changed By: W. RYAN

#1100

---

QAS: W. RYAN

#1100

GOODYEAR PD

---

Operator: W. RYAN

#1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 12/13/1959

Sex: M

Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	00:02:19
Diagnostic Test	Pass	00:02:54
Air Blank	0.000	00:03:21
0.100 Cal Check	0.097	00:03:42
Air Blank	0.000	00:04:12
Subject Test	0.045	00:04:33
Air Blank	ABT*	00:04:43
Air Blank	0.000	00:05:11

\*Sequence Aborted

Not a Successfully Completed Test Sequence



INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/14/2017

00:05:16

---

Standard Lot#: 14816100A1

Last Changed By: W. RYAN #1100

---

QAS: W. RYAN #1100

GOODYEAR PD

---

Operator: W. RYAN #1100

GOODYEAR PD

---

Subject: TEST,TEST

DOB: 12/13/1959

Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	00:06:12
Diagnostic Test	Pass	00:06:48
Air Blank	0.000	00:07:15
0.100 Cal Check	0.096	00:07:36
Air Blank	0.000	00:08:06
Subject Test	0.282*	00:11:13
Air Blank	0.000	00:11:46
Air Blank	0.000	00:12:13
Subject Test	XXX**	00:12:46
Air Blank	0.000	00:13:20

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

12/14/2017

00:14:20

-----  
Standard Lot#: 14816100A1

Last Changed By: W. RYAN #1100

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      00:14:49  
0.100 Cal Check                0.095                      00:15:10  
Air Blank                      0.000                      00:15:40

Preventative Maintenance  
Performed -- Timer Reset