

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GULMANT 1144 AGENCY GOOD YEAR PD  
DATE 4/13/2012 TIME 1807  
INTOXILYZER SERIAL # 80-000312 LOCATION 10701 S ESTRELLA Pkwy

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 102 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 10701 S ESTRELLA  
Serial Number: 80-000312  
Core Version: 8105.48  
04/13/2012

18:07:46



-----  
Standard Lot#: 06011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	18:08:16
0.100 Cal Check	0.102	18:08:37
Air Blank	0.000	18:09:07

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUTENMAUT AGENCY GOODYEAR POLICE DEPT.  
DATE 4/13/2012 TIME 1811  
INTOXILYZER SERIAL # 80-000312 LOCATION 190 N LITCHFIELD RD

- (P) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 102 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

INTOXILYZER 8000

Location: 190 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

04/13/2012

18:11:21

*ts*

-----  
Standard Lot#: 01011100A1

Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144

GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	18:11:50
0.100 Cal Check	0.102	18:12:11
Air Blank	0.000	18:12:41

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUEBASOV 114 AGENCY GOOD YEAR PD  
DATE 4/14/2012 TIME 0337  
INTOXILYZER SERIAL # 80-000312 LOCATION 190 N LETCHER RD

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution: 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 100 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 114

INTOXILYZER 8000

Location: 190 S LITCHFIELD RD

Serial Number: 80-000312

Core Version: 8105.48

04/14/2012

03:37:01

-----  
Standard Lot#: 01011100A1

Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144

GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	03:37:30
0.100 Cal Check	0.100	03:37:51
Air Blank	0.000	03:38:21

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUERRA, 1144 AGENCY GOODYEAR RA

DATE 4/18/12 TIME 1215

INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD

(\*) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

- (\*) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (\*) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (\*) 4. Air blank completed.
- (\*) 5. Calibration check completed. Test results 0. 102 AC.
- (\*) 6. Air blank completed.
- (\*) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/18/2012

12:15:08

*1144*

-----  
Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	12:15:37
0.100 Cal Check	0.102	12:15:58
Air Blank	0.000	12:16:28

Preventative Maintenance  
Performed -- Timer Reset





4/18/2012  
#  
80-000312

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

## Certificate of Analysis


Certificate ID: 2488  
Part #: BAC105L100T  
Cylinder Size: 105L  
Lot Number: 34011100A1  
Expiration: 2/1/2014

### 0.100 BAC (For use with breath alcohol testing instruments)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	260.5 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

\*NIST Standard Reference Material  
Certification of NTRM Batch No. 091602  
Nominal 210 µmol/mol Ethanol in Nitrogen  
for ILMO Products Co., Jacksonville, IL

  
Specialty Gas Lab Tech

01/09/12  
Date

Distributed by: CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)




EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUILBAULT 0144 AGENCY GOODYEAR PD  
DATE 4/18/2012 TIME 1217  
INTOXILYZER SERIAL # 80-00312 LOCATION 1111 S LITCHFIELD RD

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 101 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/18/2012

12:17:45

*M. Guilbault*  
1144

-----  
Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	12:18:15
0.100 Cal Check	0.101	12:18:36
Air Blank	0.000	12:19:06

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. FURBULT AGENCY GOODYEAR POLICE  
DATE 4/24/2012 1144 TIME 1321  
INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD.

- (2) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (3) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
4. Air blank completed.  
5. Calibration check completed. Test results 0. 100 AC.  
6. Air blank completed.  
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-00Q312  
Core Version: 8105.48  
04/24/2012

13:21:29 \*

-----  
Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	13:22:01
0.100 Cal Check	0.100	13:22:22
Air Blank	0.000	13:22:51

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUILMAULT 1144 AGENCY GOOD YEAR AD  
DATE 4/24/2012 TIME 1327  
INTOXILYZER SERIAL # 20-000312 LOCATION 1111 S LITCHFIELD RD.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE [Signature] 1144



INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/24/2012

13:27:08

---

Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144

---

QAS: M. GUILBAULT #1144  
GOODYEAR PD

---

Operator: M. GUILBAULT #1144  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:27:55
Diagnostic Test	Pass	13:28:30
Air Blank	0.000	13:28:57
0.100 Cal Check	0.101	13:29:18
Air Blank	0.000	13:29:48
Subject Test	0.000	13:30:00
Air Blank	IPS*	13:30:07
Air Blank	RFI**	13:30:15

\*Improper Sample  
\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/24/2012

13:30:20

*Handwritten initials and numbers*

Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/2101	Time
Air Blank	ABT*	13:30:49
Air Blank	0.000	13:31:18

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**



INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/24/2012

13:32:20

Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST

DOB: 01/01/1990

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:33:17
Diagnostic Test	Pass	13:33:52
Air Blank	0.000	13:34:19
0.100 Cal Check	0.101	13:34:40
Air Blank	0.000	13:35:10
Subject Test	RGE*	13:35:27
Air Blank	0.000	13:36:03

\*Range Exceeded

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/24/2012

13:37:16

Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:38:05
Diagnostic Test	Pass	13:38:40
Air Blank	0.000	13:39:07
0.100 Cal Check	0.100	13:39:28
Air Blank	0.000	13:39:58
Subject Test	RGE*	13:40:16
Air Blank	0.000	13:40:54

\*Range Exceeded

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/24/2012

13:40:58

Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST

DOB: 01/01/1990

Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:41:49
Diagnostic Test	Pass	13:42:24
Air Blank	0.000	13:42:51
0.100 Cal Check	0.100	13:43:12
Air Blank	0.000	13:43:42
Subject Test	0.042*	13:46:50
Air Blank	0.000	13:47:23
Air Blank	0.000	13:47:50
Subject Test	XXX**	13:48:23
Air Blank	0.000	13:48:56

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GILMAN LT 1144 AGENCY GOODYEAR PD  
DATE 4/24/2012 TIME 1349  
INTOXILYZER SERIAL # 80-000312 LOCATION 111 S VINTAGE RD

- (~~10~~) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
4. Air blank completed.  
5. Calibration check completed. Test results 0. 100 AC.  
6. Air blank completed.  
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
04/24/2012

13:49:07

*Handwritten: 1144*

-----  
Standard Lot#: 34011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	13:49:36
0.100 Cal Check	0.100	13:49:57
Air Blank	0.000	13:50:27

Preventative Maintenance  
Performed -- Timer Reset