

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUELGAULT 1144 AGENCY GOODYEAR PD
DATE 12/1/2009 TIME 12:14:29
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PA

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 104 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  1144


INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-000312
12/01/2009

14:29:14

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:29:41
Cal Check	0.104	14:30:00
Air Blank	0.000	14:30:30



Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GURJAGULT 1144 AGENCY GOODYEAR PD
DATE 12/11/2009 TIME 1430
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 104 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number : 80-000312
12/01/2009

14:30:37

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST, TEST
DOB: 01/01/2000
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:31:29
Diagnostic Test	Pass	14:31:56
Air Blank	0.000	14:32:23
0.100 Cal Check	0.105	14:32:42
Air Blank	0.000	14:33:12
Subject Test	0.000	14:33:27
Air Blank	ABT*	14:33:39
Air Blank	0.000	14:34:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number : 80-000312
12/01/2009

14:34:44

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST, TEST
DOB: 01/01/2000
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	14:35:43
Air Blank	RFI**	14:35:51

*Improper Sample *
**RFI Detect *

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number : 80-000312
12/01/2009

14:35:56

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST, TEST
DOB: 01/01/2000
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:36:48
Diagnostic Test	Pass	14:37:15
Air Blank	0.000	14:37:42
0.100 Cal Check	0.104	14:38:01
Air Blank	0.000	14:38:31
Subject Test	0.015*	14:41:39
Air Blank	0.000	14:42:11
Air Blank	0.000	14:42:38
Subject Test	XXX**	14:42:58
Air Blank	0.000	14:43:33

~~*Deficient Sample~~

~~**Invalid Sample~~

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUEVAULT 1144 AGENCY GOODYEAR PD
DATE 12/1/2009 TIME 1444
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 103 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1144

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-000312
12/01/2009

14:44:24

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:44:51
Cal Check	0.103	14:45:10
Air Blank	0.000	14:45:39


Operator's Signature


INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-000312
12/29/2009

16:01:13

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	16:01:41
Cal Check	0.104	16:02:00
Air Blank	0.000	16:02:29



Operator's Signature
1144