

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUSMANOV 1144 AGENCY GOODYEAR PD
DATE 7/5/11 TIME 1701
INTOXILYZER SERIAL # 80-000312 LOCATION 10701 SESTRELLA PKWY

gk 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- gk* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 100 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____
M. Gusmanov

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
07/05/2011

17:01:58

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	17:02:32
0.100 Cal Check	0.100	17:02:53
Air Blank	0.000	17:03:22

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. G. VIELMANT 1144 AGENCY GOOD YEAR PD
DATE 7/5/11 TIME 1704
INTOXILYZER SERIAL # 80-000312 LOCATION 10701 SESTRELLA LANE

() T. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE  1144

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
07/05/2011

17:04:20

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:05:08
Diagnostic Test	Pass	17:05:43
Air Blank	0.000	17:06:10
0.100 Cal Check	0.099	17:06:31
Air Blank	0.000	17:07:01
Subject Test	0.000	17:07:15
Air Blank	0.000	17:07:42
Wait	ABT*	
Air Blank	RFI**	17:07:56

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
07/05/2011

17:08:01

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	17:08:30
Air Blank	0.000	17:08:58

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
07/05/2011

17:09:01

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:09:52
Diagnostic Test	Pass	17:10:26
Air Blank	0.000	17:10:54
0.100 Cal Check	0.100	17:11:15
Air Blank	0.000	17:11:45
Subject Test	0.375*	17:14:52
Air Blank	0.000	17:15:27
Air Blank	0.000	17:15:54
Subject Test	XXX**	17:16:23
Air Blank	0.000	17:16:57

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUERBAULT 1144 AGENCY GOODSHEAR PD
DATE 7/5/11 TIME 1717
INTOXILYZER SERIAL # 80-000312 LOCATION 10701 S ESTANOLA AVE 7

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
07/05/2011

17:17:09

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	17:17:38
0.100 Cal Check	0.099	17:17:59
Air Blank	0.000	17:18:29

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. [Signature] 1144 AGENCY GOODYEAR PD
DATE 7/26/2011 TIME 0132
INTOXILYZER SERIAL # BD-000312 LOCATION 10701 S ESTRELLA PKWY

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.094 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1144

INTOXILYZER 8000

Location: 10701 S ESTRELLA

Serial Number: 80-000312

Core Version: 8105.48

07/06/2011

01:32:00

Standard Lot#: 01011100A1

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 01:32:57
0.100 Cal Check 0.094 01:33:18
Air Blank 0.000 01:33:48

Preventative Maintenance
Performed -- Timer Reset