

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. LEVELBAULT AGENCY GOODYEAR PD
DATE 6/16/2010 TIME 1257
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD (VENEDA)

- (A) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (A) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (A) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (A) 4. Air blank completed.
- (A) 5. Calibration check completed. Test results 0. 099 AC.
- (A) 6. Air blank completed.
- (A) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____ 1144

INTOXILYZER 8000

Location: GOODYEAR PD VENIDA

Serial Number: 80-000312

Core Version: 8105.48

06/16/2010

12:57:25

Standard Lot#: 672733

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 12:57:55
0.100 Cal Check 0.099 12:58:16
Air Blank 0.000 12:58:46

JK
Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUEUBAULT AGENCY GOODYEAR PD
DATE 6/16/2010 TIME 1259
INTOXILYZER SERIAL # 96-000 J12 LOCATION GOODYEAR PD (VENZIA)

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE [Signature] 1144

INTOXILYZER 8000
Location: GOODYEAR PD VENIDA
Serial Number: 80-000312
Core Version: 8105.48
06/16/2010

12:59:30

Standard Lot#: 672733
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:00:58
Diagnostic Test	Pass	13:01:33
Air Blank	0.000	13:02:00
0.100 Cal Check	0.100	13:02:21
Air Blank	0.000	13:02:51
Subject Test	0.000	13:03:05
Air Blank	0.000	13:03:32
Wait	ABT*	
Air Blank	0.000	13:04:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD VENIDA
Serial Number: 80-000312
Core Version: 8105.48
06/16/2010

13:04:13

Standard Lot#: 672733
Last Changed By: M. GUILBAULT #1144

OAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	13:04:43
Air Blank	0.000	13:05:12

*Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD VENIDA
Serial Number: 80-000312
Core Version: 8105.48
06/16/2010

13:05:17

Standard Lot#: 672733
Last Changed By: M. GUILBAULT #1144

CAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	13:05:46
Air Blank	0.000	13:06:15

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD VENIDA
Serial Number: 80-000312
Core Version: 8105.48
06/16/2010

13:06:28

Standard Lot#: 672733
Last Changed By: M. GUILBAULT #1144
OAS: M. GUILBAULT #1144
GOODYEAR PD
Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:07:18
Diagnostic Test	Pass	13:07:53
Air Blank	0.000	13:08:21
0.100 Cal Check	0.099	13:08:41
Air Blank	0.000	13:09:12
Subject Test	0.024*	13:12:19
Air Blank	0.000	13:12:51
Air Blank	0.000	13:13:18
Subject Test	XXX**	13:13:35
Air Blank	0.000	13:14:09

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUERBAU AGENCY GOODYEAR PD

DATE 6/16/2010 TIME _____

INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD (VENUE)

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10% AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 100 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1144

INTOXILYZER 8000

Location: GOODYEAR PD VENIDA

Serial Number: 80-000312

Core Version: 8105.48

06/16/2010

13:14:27

Standard Lot#: 672733

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test g/210L Time

Air Blank 0.000 13:14:57
0.100 Cal Check 0.100 13:15:18
Air Blank 0.000 13:15:47

Preventative Maintenance
Performed -- Timer Reset



EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE


QA SPECIALIST M-GUEMBALT AGENCY GOODYEAR PD
DATE 6/24/2010 TIME 2129
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD VENETA

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0.102 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____


1144

INTOXILYZER 8000
Location: GOODYEAR PD VENIDA
Serial Number: 80-000312
Core Version: 8105.48
06/24/2010

21:29:30

Standard Lot#: 672733
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:30:08
0.100 Cal Check	0.102	21:30:29
Air Blank	0.000	21:30:59

Preventative Maintenance
Performed -- Timer Reset

