

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W. T. RYAN #1100 AGENCY GOODYEAR POLICE DEPT
DATE 3-2-2019 TIME 1644 HRS
INTOXILYZER SERIAL # 80-001858 LOCATION 11 N 145TH AVE, GOODYEAR AZ 85338

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W. T. Ryan #1100

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-001858

Core Version: 8105.48

03/02/2019

16:44:57

Standard Lot#: 10818100A1

Last Changed By: W. RYAN #1100

QAS: w. ryan #1100

goodyear pd

Operator: w. ryan #1100

goodyear pd

Test	g/210L	Time
Air Blank	0.000	16:45:29
0.100 Cal Check	0.100	16:45:50
Air Blank	0.000	16:46:20

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR POLICE DEPT
DATE 3-2-2019 TIME 1647 HRS
INTOXILYZER SERIAL # 80-001858 LOCATION 11 N 145TH AVE, GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START" 85338

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE W.T. Ryan #1100

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-001858
Core Version: 8105.48
03/02/2019

16:47:59

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

QAS: w. ryan #1100
goodyear pd

Operator: w. ryan #1100
goodyear pd

Subject: TEST,TEST
DOB: 03/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:49:03
Diagnostic Test	Pass	16:49:38
Air Blank	0.000	16:50:05
0.100 Cal Check	0.100	16:50:26
Air Blank	0.000	16:50:57
Subject Test	0.000	16:51:11
Air Blank	0.000	16:51:38
Wait	ABT*	
Air Blank	0.000	16:52:28

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-001858

Core Version: 8105.48

03/02/2019

16:52:46

Standard Lot#: 10818100A1

Last Changed By: W. RYAN

#1100

QAS: w. ryan

#1100

goodyear pd

Operator: w. ryan

#1100

goodyear pd

Subject: TEST,TEST

DOB: 03/02/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:53:24
Air Blank	0.000	16:53:52

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-001858

Core Version: 8105.48

03/02/2019

16:54:38

Standard Lot#: 10818100A1

Last Changed By: W. RYAN #1100

QAS: w. ryan #1100
goodyear pd

Operator: w. ryan #1100
goodyear pd

Subject: TEST,TEST

DOB: 03/02/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:55:15
Air Blank	0.000	16:55:43

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-001858

Core Version: 8105.48

03/02/2019

16:56:37

Standard Lot#: 10818100A1

Last Changed By: W. RYAN #1100

QAS: w. ryan #1100
goodyear pd

Operator: w. ryan #1100
goodyear pd

Subject: TEST,TEST

DOB: 03/02/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:57:33
Diagnostic Test	Pass	16:58:08
Air Blank	0.000	16:58:36
0.100 Cal Check	0.100	16:58:57
Air Blank	0.000	16:59:27
Subject Test	0.295*	17:02:34
Air Blank	0.000	17:03:08
Air Blank	0.000	17:03:35
Subject Test	XXX**	17:04:01
Air Blank	0.000	17:04:36

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-001858

Core Version: 8105.48

03/02/2019

17:05:37

Standard Lot#: 10818100A1

Last Changed By: W. RYAN #1100

QAS: w. ryan #1100

goodyear pd

Operator: w. ryan #1100

goodyear pd

Test	g/210L	Time
Air Blank	0.000	17:06:07
0.100 Cal Check	0.098	17:06:27
Air Blank	0.000	17:06:58

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000


STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. Guibault AGENCY GPD
DATE 3/29/19 TIME 1215
INTOXILYZER SERIAL # 80-001858 LOCATION GPD

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE GA 1144

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-001858
Core Version: 8105.48
03/29/2019

12:15:10 

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

QAS: m. guilbault #1144
goodyear pd

Operator: m. guilbault #1144
goodyear pd

Test	g/210L	Time
Air Blank	0.000	12:15:40
0.100 Cal Check	0.101	12:16:01
Air Blank	0.000	12:16:31

Preventative Maintenance
Performed -- Timer Reset



Canister change
MG
3/29/19

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 11089
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 10818100A1
Expiration: 6/5/2020

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	260 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:


CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*NIST Traceable Reference Material
Cylinder No. CC274523 / Job No. 09160306
Certified 362.2 µmol/mol Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

05-07-18
Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000


STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. Gohlhauert AGENCY GOODYEAR PD
DATE 3/29/19 TIME 1217
INTOXILYZER SERIAL # 80-061858 LOCATION GPD

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 101 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE AA 1144

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-001858
Core Version: 8105.48
03/29/2019

12:17:47 

Standard Lot#: 10818100A1
Last Changed By: m. guilbault #1144

QAS: m. guilbault #1144
goodyear pd

Operator: m. guilbault #1144
goodyear pd

Test	g/210L	Time
Air Blank	0.000	12:18:16
0.100 Cal Check	0.101	12:18:37
Air Blank	0.000	12:19:08

Preventative Maintenance
Performed -- Timer Reset