

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. LOUBAULT 1144 AGENCY GOODYEAR PD
DATE 11/12/2010 TIME 1249
INTOXILYZER SERIAL # FD-000312 LOCATION GAD #2

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 103 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
11/12/2010

12:49:55

Standard Lot#: 767153
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD



Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	12:50:27
0.100 Cal Check	0.103	12:50:47
Air Blank	0.000	12:51:17

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUELBAULT 1144 AGENCY GOODYEAR PD
DATE 11/17/2010 TIME 1419
INTOXILYZER SERIAL # 80-000312 LOCATION GPA #2

- ~~1~~) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ~~2~~) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ~~3~~) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ~~4~~) 4. Air blank completed.
- ~~5~~) 5. Calibration check completed. Test results 0. 102 AC.
- ~~6~~) 6. Air blank completed.
- ~~7~~) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
11/17/2010

14:19:34

Standard Lot#: 767153
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:20:14
0.100 Cal Check	0.102	14:20:35
Air Blank	0.000	14:21:05

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUELBAULT 1144 AGENCY GOODYEAR POLICE DEPT.
DATE 11/17/2010 TIME 1421
INTOXILYZER SERIAL # 80-000312 LOCATION GPD #2

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

1144

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
11/17/2010

14:21:22

Standard Lot#: 767153
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:22:18
Diagnostic Test	Pass	14:22:53
Air Blank	0.000	14:23:20
0.100 Cal Check	0.102	14:23:41
Air Blank	0.000	14:24:11
Subject Test	0.000	14:24:32
Air Blank	IPS*	14:24:39
Air Blank	RFI**	14:24:55

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
11/17/2010

14:25:06

Standard Lot#: 767153
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:25:55
Diagnostic Test	Pass	14:26:30
Air Blank	0.000	14:26:57
0.100 Cal Check	0.102	14:27:18
Air Blank	0.000	14:27:48
Subject Test	0.000	14:28:01
Air Blank	ABT*	14:28:12
Air Blank	0.000	14:28:41

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
11/17/2010

14:28:46

Standard Lot#: 767153
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:29:36
Diagnostic Test	Pass	14:30:11
Air Blank	0.000	14:30:38
0.100 Cal Check	0.102	14:30:59
Air Blank	0.000	14:31:28
Subject Test	0.201*	14:34:36
Air Blank	0.000	14:35:08
Air Blank	0.000	14:35:36
Subject Test	XXX**	14:36:03
Air Blank	0.000	14:36:38

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE


QA SPECIALIST M. GUILBAULT 1144 AGENCY GOODYEAR PD
DATE 11/17/2010 TIME 1437
INTOXILYZER SERIAL # 80-000312 LOCATION GAD #2

- ~~1~~ 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ~~2~~ 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ~~3~~ 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ~~4~~ 4. Air blank completed.
- ~~5~~ 5. Calibration check completed. Test results 0. 101 AC.
- ~~6~~ 6. Air blank completed.
- ~~7~~ 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-000312
Core Version: 8105.48
11/17/2010

14:37:15



Standard Lot#: 767153
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:37:44
0.100 Cal Check	0.101	14:38:05
Air Blank	0.000	14:38:35

Preventative Maintenance
Performed -- Timer Reset