



ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MAINTENANCE REPORT

AGENCY Goodyear Police Department
Goodyear, AZ 85338
OFFICER WILL, #974
DATE November 08, 2013

Page 1 of 1

INSTRUMENT:

80000312

MAINTENANCE REQUESTED:

Repair and Preventative Maintenance

MAINTENANCE PERFORMED:

Parts Replaced:

Breath Hose

Preventative Maintenance Performed:

Inspection, Calibration, Verification, Adjustment

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

JOSEPH M SLOWINSKI, #6101
Electronics Technician
Central Regional Crime Laboratory
P. O. Box 6638
Phoenix, AZ 85005
(602) 223-2394



EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR POLICE DEPT
DATE 11-13-13 TIME 1749
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR POLICE DEPT

(☒) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (☒) 1. Clock time check.
(☒) 2. Date check.

OPERATIONAL TESTS

- (☒) 1. Alcohol-free subject test result 0. 000 AC.
(☒) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(☒) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(☒) 4. Standard Calibration Check standard 0. 10 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS

JUST RETURNED FROM DPS

SIGNATURE

W.T. Ryan 1100

INTOXILYZER 8000 .
Location: 10701 S ESTRELLA PKW
Serial Number: 80-000312
Core Version: 8105.48
11/13/2013 17:49:23

Standard Lot#: 32012100A1
Last Changed By: M. GUILBAULT #1144

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 11/13/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:50:24
Diagnostic Test	Pass	17:50:59
Air Blank	0.000	17:51:26
0.100 Cal Check	0.099	17:51:47
Air Blank	0.000	17:52:17
Subject Test	0.000	17:52:36
Air Blank	0.000	17:53:03
Wait	ABT*	
Air Blank	0.000	17:55:17

*Sequence Aborted .

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA PKW
Serial Number: 80-000312
Core Version: 8105.48
11/13/2013 17:56:44

Standard Lot#: 32012100A1
Last Changed By: M. GUILBAULT #1144

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 11/13/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	17:57:29
Air Blank	0.000	17:57:58

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 10701 S ESTRELLA PKW

Serial Number: 80-000312

Core Version: 8105.48

11/13/2013

17:58:34

Standard Lot#: 32012100A1

Last Changed By: M. GUILBAULT #1144

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/13/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	17:59:36
Air Blank	0.000	18:00:04

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 10701 S ESTRELLA PKW

Serial Number: 80-000312

Core Version: 8105.48

11/13/2013

18:00:54

Standard Lot#: 32012100A1

Last Changed By: M. GUILBAULT #1144

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/13/1959

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:02:06
Diagnostic Test	Pass	18:02:41
Air Blank	0.000	18:03:08
0.100 Cal Check	0.098	18:03:29
Air Blank	0.000	18:03:59
Subject Test	0.048*	18:07:06
Air Blank	0.000	18:07:39
Air Blank	0.000	18:08:06
Subject Test	RGE**	18:08:20
Air Blank	0.000	18:09:06

*Deficient Sample

**Range Exceeded

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 10701 S ESTRELLA PKW

Serial Number: 80-000312

Core Version: 8105.48

11/13/2013

18:09:15

Standard Lot#: 32012100A1

Last Changed By: M. GUILBAULT #1144

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Subject: TEST,TEST

DOB: 11/13/1959

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:10:09
Diagnostic Test	Pass	18:10:44
Air Blank	0.000	18:11:11
0.100 Cal Check	0.098	18:11:32
Air Blank	0.000	18:12:02
Subject Test	0.036*	18:15:09
Air Blank	0.000	18:15:41
Air Blank	0.000	18:16:09
Subject Test	XXX**	18:16:33
Air Blank	0.000	18:17:15

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 10701 S ESTRELLA PKW

Serial Number: 80-000312

Core Version: 8105.48

11/13/2013

18:19:27

Standard Lot#: 32012100A1

Last Changed By: M. GUILBAULT #1144

QAS: W. RYAN #1100

GOODYEAR PD

Operator: W. RYAN #1100

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	18:19:58
0.100 Cal Check	0.098	18:20:18
Air Blank	0.000	18:20:49

Preventative Maintenance

Performed -- Timer Reset