

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Will AGENCY Goodyear PD  
DATE 10/07/2009 TIME 1708  
INTOXILYZER SERIAL # 80-00012 LOCATION CD

*W* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- W* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
*W* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
*W* 4. Air blank completed.  
*W* 5. Calibration check completed. Test results 0. 103 AC.  
*W* 6. Air blank completed.  
*W* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *Will*

INTOXILYZER 8000

Location: GOODYEAR PD

Serial Number: 80-000312

10/07/2009

17:08:25


-----  
Standard Lot#: 562576

Last Changed By: M. GUILBAULT #1144  
-----

QAS: W. WILL #974

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	17:08:52
Cal Check	0.103	17:09:11
Air Blank	0.000	17:09:40

  
-----  
Operator's Signature