

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. COLEBALT 1144 AGENCY GOOD YEAR PD
DATE 9/5/2012 TIME 0421
INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD

- (*) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (*) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (*) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (*) 4. Air blank completed.
- (*) 5. Calibration check completed. Test results 0. 098 AC.
- (*) 6. Air blank completed.
- (*) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1144

INTOXILYZER 8000
Location: 1111 S LICHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
09/05/2012

04:21:54

Standard Lot#: 34011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:22:26
0.100 Cal Check	0.098	04:22:47
Air Blank	0.000	04:23:17

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. G. WILSON 1144 AGENCY GOODYEAR PD
DATE 9/5/2012 TIME 0423
INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD RD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

INTOXILYZER 8000
Location: 1111 S LICHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
09/05/2012

04:23:25

Standard Lot#: 34011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:24:11
Diagnostic Test	Pass	04:24:46
Air Blank	0.000	04:25:14
0.100 Cal Check	0.098	04:25:35
Air Blank	0.000	04:26:04
Subject Test	0.000	04:26:17
Air Blank	RFI*	04:26:29
Air Blank	IPS**	04:26:39

*RFI Detect
**Imple

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LICHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
09/05/2012

04:26:44

Standard Lot#: 34011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:27:30
Diagnostic Test	Pass	04:28:06
Air Blank	ABT*	04:28:29
Air Blank	0.000	04:28:58

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LICHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
09/05/2012

04:29:02



Standard Lot#: 34011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:29:48
Diagnostic Test	Pass	04:30:24
Air Blank	0.000	04:30:51
0.100 Cal Check	0.098	04:31:12
Air Blank	0.000	04:31:42
Subject Test	0.081*	04:34:49
Air Blank	0.000	04:35:22
Air Blank	0.000	04:35:49
Subject Test	XXX**	04:36:06
Air Blank	0.000	04:36:39

*Deficient Sample
**Invalid Sample
Begin new deprivation period


Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUEMOUNT 1144 AGENCY GOODYEAR PD
DATE 9/5/2012 TIME 0436
INTOXILYZER SERIAL # 80-000312 LOCATION 1111 S LITCHFIELD

- (A) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (A) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (A) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (A) 4. Air blank completed.
- (A) 5. Calibration check completed. Test results 0. 098 AC.
- (A) 6. Air blank completed.
- (A) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000
Location: 1111 S LICHFIELD RD
Serial Number: 80-000312
Core Version: 8105.48
09/05/2012

04:36:58

A

Standard Lot#: 34011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:37:28
0.100 Cal Check	0.098	04:37:49
Air Blank	0.000	04:38:19

Preventative Maintenance
Performed -- Timer Reset