

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W. T. Ryan 1100 AGENCY GOODYEAR POLICE DEPT

DATE 8-1-2019 TIME 1834 HRS

INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE GOODYEAR AZ 85338

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W. T. Ryan 1100

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
08/01/2019

18:34:28

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: W. RYAN #1100  
GOODYEAR PD

Operator: W. RYAN #1100  
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	18:35:16
0.100 Cal Check	0.097	18:35:37
Air Blank	0.000	18:36:07

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W. T. Ryan 1100 AGENCY GOODYEAR POLICE DEPT  
DATE 8-1-2019 TIME 1841 HRS  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE GOODYEAR AZ 85338

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE W. T. Ryan 1100

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
08/01/2019 18:41:35

-----  
Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100  
-----

QAS: W. RYAN #1100  
GOODYEAR PD  
-----

Operator: W. RYAN #1100  
GOODYEAR PD  
-----

Subject: TEST,TEST  
DOB: 08/01/1959  
Sex: M Weight: 185  
-----

15 Minute Deprivation Period? Yes

-----  
Test g/210L Time  
-----  
Air Blank 0.000 18:42:32  
Diagnostic Test Pass 18:43:09  
Air Blank 0.000 18:43:37  
0.100 Cal Check 0.097 18:43:57  
Air Blank 0.000 18:44:27  
Subject Test 0.000 18:44:41  
Air Blank 0.000 18:45:09  
Wait ABT\*  
Air Blank 0.000 18:45:43

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 6105.48  
08/01/2019 18:47:25

-----  
Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100  
-----

QAS: W. RYAN #1100  
GOODYEAR PD  
-----

Operator: W. RYAN #1100  
GOODYEAR PD  
-----

Subject: TEST,TEST  
DOB: 08/01/1959  
Sex: M Weight: 185  
-----

15 Minute Deprivation Period? Yes  
-----

Test	gr/210L	Time
Air Blank	1PS*	18:47:59
Air Blank	0.000	18:48:28

\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
08/01/2019 18:49:35

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: W. RYAN #1100  
GOODYEAR PD

Operator: W. RYAN #1100  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 08/01/1959  
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	18:50:10
Air Blank	0.000	18:50:39

\*RFI Detect

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
08/01/2019 18:51:32

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: W. RYAN #1100  
GOODYEAR PD

Operator: W. RYAN #1100  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 08/01/1959  
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:52:23
Diagnostic Test	Pass	18:53:01
Air Blank	0.000	18:53:28
0.100 Cal Check	0.098	18:53:49
Air Blank	0.000	18:54:19
Subject Test	0.058*	18:57:27
Air Blank	0.000	18:57:58
Air Blank	0.000	18:58:25
Subject Test	XXX**	18:59:20
Air Blank	0.000	18:59:54

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
08/01/2019 19:00:53

-----  
Standard Lot#: 10818100A!  
Last Changed By: W. RYAN #1100  
-----

DAS: W. RYAN #1100  
GOODYEAR PD  
-----

Operator: W. RYAN #1100  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	19:01:22
0.100 Cal Check	0.096	19:01:43
Air Blank	0.000	19:02:14

Preventative Maintenance  
Performed -- Timer Reset