

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUELMAUT AGENCY GOODYEAR PA
DATE 9/9/09 TIME 1035
INTOXILYZER SERIAL # 80000312 LOCATION GOODYEAR #2

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 103 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-000312
09/09/2009

10:34:43

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	10:35:10
Cal Check	0.103	10:35:29
Air Blank	0.000	10:35:58



Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUDENRUT AGENCY GOODYEAR PD
DATE 9/9/09 TIME 1037
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD #2

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE [Signature]

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number : 80-000312
09/09/2009

10:36:29

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:37:49
Diagnostic Test	Pass	10:38:16
Air Blank	0.000	10:38:44
0.100 Cal Check	0.102	10:39:02
Air Blank	0.000	10:39:32
Subject Test	0.000	10:39:49
Air Blank	0.000	10:40:16
Five Minute Wait	ABT*	
Air Blank	0.000	10:40:52

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-000312
09/05/2009 10:41:25

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

OAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: SOAP, TEST
DOB: 01/01/2000
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:42:11
Air Blank	RFI**	10:42:26

*Improper Sample
**RFI Detect

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number : 80-000312
09/09/2009

10:42:47

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: SQAP, TEST

DOB: 01/01/2000

Sex: M

Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:43:47
Diagnostic Test	Pass	10:44:16
Air Blank	0.000	10:44:43
0.100 Cal Check	0.102	10:45:02
Air Blank	0.000	10:45:31
Subject Test	0.064*	10:48:40
Air Blank	0.000	10:49:14
Air Blank	0.000	10:49:41
Subject Test	XXX**	10:50:17
Air Blank	0.000	10:50:51

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. Guebert AGENCY GOODYEAR PD
DATE 9/9/09 TIME 1052
INTOXILYZER SERIAL # 80-000312 LOCATION GOODYEAR PD #2

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]


INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-000312
09/09/2009

10:51:34

Standard Lot#: 562576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	10:52:01
Cal Check	0.101	10:52:20
Air Blank	0.000	10:52:49


Operator's Signature