

Memo



To: Cal. Records Intox – 80-001858
From: Officer M. Guilbault #1144
Date: 8/5/2010
Re: Intox 80-001858

Upon receiving Intox 80-001858 back from DPS for software update the instrument timed out and locked out the QAS Users, Officer Guilbault and Officer Will. Intox 80-001858 was taken back to DPS of which Officer Will was instructed of how to unlock the instrument with the new software update, power down and then power back up.

After this a seal gasket was broken on the gas tank of which we obtained a new rubber seal thus allowing proper use and cal checks to take place.

Therefore there is no reported data for the month of April.