

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR P.D.

DATE 4-26-14 TIME 1619 HRS

INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD GOODYEAR AZ

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE W.T. Ryan 1100

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
04/26/2014 16:19:45

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 04/26/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:21:28
Diagnostic Test	Pass	16:22:03
Air Blank	0.000	16:22:31
0.100 Cal Check	0.098	16:22:52
Air Blank	0.000	16:23:21
Subject Test	0.000	16:23:37
Air Blank	0.000	16:24:04
Wait	ABT*	
Air Blank	0.000	16:24:39

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
04/26/2014 16:25:24

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 04/26/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:26:19
Air Blank	0.000	16:26:47

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
04/26/2014 16:27:36

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 04/26/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:28:20
Air Blank	RFI*	16:28:32

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
04/26/2014 16:30:22

---

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

---

QAS: W. RYAN #1100  
GOODYEAR PD

---

Operator: W. RYAN #1100  
GOODYEAR PD

---

Subject: TEST,TEST  
DOB: 04/26/1959  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:31:15
Diagnostic Test	Pass	16:31:50
Air Blank	0.000	16:32:18
0.100 Cal Check	0.098	16:32:39
Air Blank	0.000	16:33:08
Subject Test	0.031*	16:36:16
Air Blank	0.000	16:36:47
Air Blank	0.000	16:37:15
Subject Test	XXX**	16:37:32
Air Blank	0.000	16:38:08

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

04/26/2014

16:39:29

-----  
Standard Lot#: 32012100A1

Last Changed By: W. WILL #974  
-----

QAS: W. RYAN #1100

GOODYEAR PD  
-----

Operator: W. RYAN #1100

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	16:40:00
0.100 Cal Check	0.098	16:40:21
Air Blank	0.000	16:40:50

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

04/26/2014

16:41:00

-----  
Standard Lot#: 32012100A1

Last Changed By: W. WILL #974

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	16:41:55
0.100 Cal Check	0.098	16:42:15
Air Blank	0.000	16:42:45

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN #1100 AGENCY GOODYEAR P.D.  
DATE 4-26-14 TIME 1616 HRS

INTOXILYZER SERIAL # 80-001858 LOCATION 111 S LITCHFIELD RD GOODYEAR AZ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.10 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan 1100



INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

04/26/2014

16:16:31

-----  
Standard Lot#: 32012100A1

Last Changed By: W. WILL #974  
-----

QAS: W. RYAN #1100

GOODYEAR PD  
-----

Operator: W. RYAN #1100

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	16:17:07
0.100 Cal Check	0.098	16:17:28
Air Blank	0.000	16:17:57

Preventative Maintenance  
Performed -- Timer Reset