

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. ROGERS AGENCY GOODYEAR PD
DATE 8-14-08 TIME 1450
INTOXILYZER SERIAL # 80-001858 LOCATION GOODYEAR PD

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Rogers #1123

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

08/14/2008

14:50:16

Standard Lot#: 808401G

Last Changed By: W. WILL #974

QAS: J. ROGERS #1123

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:50:43
Cal Check	0.100	14:51:02
Air Blank	0.000	14:51:32



Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. ROGERS AGENCY GOODYEAR PD
DATE 8-14-08 TIME 1452
INTOXILYZER SERIAL # 80-001858 LOCATION GOODYEAR PD

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

J. Rogers #1123

INTOXILYZER 8000
Location: GOODYEAR POLICE DEPT
Serial Number : 80-001858
08/14/2008 14:52:33

Standard Lot#: 808401G
Last Changed By: W. WILL #974

QAS: J. ROGERS #1123
GOODYEAR PD

Operator: J. ROGERS #1123
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/0101
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:53:27
Diagnostic Test	Pass	14:53:55
Air Blank	0.000	14:54:22
0.100 Cal Check	0.100	14:54:41
Air Blank	0.000	14:55:11
Subject Test	0.000	14:55:30
Air Blank	0.000	14:55:58
Five Minute Wait	ABT*	
Air Blank	0.000	14:56:58

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number : 80-001858

08/14/2008

14:57:44

Standard Lot#: 808401G

Last Changed By: W. WILL

#974

QAS: J. ROGERS

#1123

GOODYEAR PD

Operator: J. ROGERS

#1123

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/01/0101

Sex: M

Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	14:58:13
Air Blank	0.000	14:58:41

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR POLICE DEPT
Serial Number : 80-001858
08/14/2008 15:07:53

Standard Lot#: 808401G
Last Changed By: W. WILL #974

QAS: J. ROGERS #1123
GOODYEAR PD

Operator: J. ROGERS #1123
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/0101
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:08:27
Air Blank	0.000	15:08:56

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number : 80-001858

08/14/2008

15:09:35

Standard Lot#: 808401G

Last Changed By: W. WILL

#974

QAS: J. ROGERS

#1123

GOODYEAR PD

Operator: J. ROGERS

#1123

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/01/0101

Sex: M

Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:10:25
Diagnostic Test	Pass	15:10:52
Air Blank	0.000	15:11:19
0.100 Cal Check	0.100	15:11:38
Air Blank	0.000	15:12:08
Subject Test	0.108*	15:15:16
Air Blank	0.000	15:15:48
Air Blank	0.000	15:16:15
Subject Test	RGE**	15:16:36
Air Blank	0.000	15:17:11

*Deficient Sample

**Range Exceeded

BLEW TOO QUICKLY AFTER
PLACING MOUTH ALCOHOL
IN MOUTH

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: GOODYEAR POLICE DEPT
Serial Number : 80-001858
08/14/2008 15:17:17

Standard Lot#: 808401G
Last Changed By: W. WILL #974

QAS: J. ROGERS #1123
GOODYEAR PD

Operator: J. ROGERS #1123
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/0101
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:18:11
Diagnostic Test	Pass	15:18:38
Air Blank	0.000	15:19:05
0.100 Cal Check	0.100	15:19:24
Air Blank	0.000	15:19:54
Subject Test	XXX*	15:20:29
Air Blank	0.000	15:21:03

*Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. ROGERS AGENCY GOODYEAR PD
DATE 8-11-08 TIME 1521
INTOXILYZER SERIAL # 80-001858 LOCATION GOODYEAR PD

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] # 1123

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

08/14/2008

15:21:42


Standard Lot#: 808401G

Last Changed By: W. WILL #974

QAS: J. ROGERS #1123

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	15:22:09
Cal Check	0.099	15:22:28
Air Blank	0.000	15:22:58

 #1123

Operator's Signature