

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR POLICE DEPT

DATE 11-2-19 TIME 0309 HRS

INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE, GOODYEAR AZ 85338

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.106 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 60-000312
Core Version: 8105.48
11/02/2019

03:09:28

Standard Lot#: 1081810091
Last Changed By: W. RYAN #1100

DAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:09:59
0.100 Cal Check	0.098	03:10:20
Air Blank	0.000	03:10:50

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR POLICE DEPT
DATE 11-2-2019 TIME 0313 HRS
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE GOODYEAR AZ 85338

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE W.T. Ryan 1100

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-000312
Core Version: 8105.48
11/02/2019 03:13:14

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

OAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST, TEST
DOB: 11/02/1959
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	03:14:13
Diagnostic Test	Pass	03:14:50
Air Blank	0.000	03:15:17
0.100 Cal Check	0.099	03:15:38
Air Blank	0.000	03:16:08
Subject Test	0.000	03:16:26
Air Blank	0.000	03:16:54
Wait	ABT*	
Air Blank	0.000	03:17:37

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 60-000312
Core Version: 8105.48
11/02/2019 03:19:17

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

DAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 11/02/1959
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test g/210L Time

Air Blank 1PS* 03:20:09
Air Blank 0.000 03:20:37

*Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-000312
Core Version: 8105.48
11/02/2019 03:22:18

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST,TEST
DOB: 11/02/1959
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	03:22:54
Air Blank	0.000	03:23:22

*RFI Detect

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-000312
Core Version: 8105.48
11/02/2019

03:24:35

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

DAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Subject: TEST, TEST
DOB: 11/02/1959
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	03:25:34
Diagnostic Test	Pass	03:26:11
Air Blank	0.000	03:26:38
0.100 Cal Check	0.098	03:26:59
Air Blank	0.000	03:27:29
Subject Test	0.062*	03:30:37
Air Blank	0.000	03:31:08
Air Blank	0.000	03:31:36
Subject Test	XXX**	03:32:17
Air Blank	0.000	03:32:50

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: 11 N 145TH AVE
Serial Number: 80-000312
Core Version: 8105.48
11/02/2019

03:33:35

Standard Lot#: 10818100A1
Last Changed By: W. RYAN #1100

QAS: W. RYAN #1100
GOODYEAR PD

Operator: W. RYAN #1100
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:34:05
0.100 Cal Check	0.097	03:34:26
Air Blank	0.000	03:34:56

Preventative Maintenance
Performed -- Timer Reset