

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GURBAULT AGENCY GOODYEAR POLICE DEPT.  
DATE 21 6/2012 TIME 1814  
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD

- ( ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ( ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ) 4. Air blank completed.
- ( ) 5. Calibration check completed. Test results 0. 100 AC.
- ( ) 6. Air blank completed.
- ( ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
02/06/2012

18:14:34

A

-----  
Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	18:15:04
0.100 Cal Check	0.100	18:15:25
Air Blank	0.000	18:15:54

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUILBAULT AGENCY GOODYEAR PD  
DATE 2/6/2012 TIME 1816  
INTOXILYZER SERIAL # 80001858 LOCATION 111 S LITCHFIELD RD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE [Signature] 1144

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
02/06/2012

18:16:17

Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:17:03
Diagnostic Test	Pass	18:17:38
Air Blank	0.000	18:18:06
0.100 Cal Check	0.100	18:18:26
Air Blank	0.000	18:18:56
Subject Test	0.000	18:19:11
Air Blank	IPS*	18:19:17
Air Blank	RFI**	18:19:25

\*Improper Sample

\*\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
02/06/2012

18:19:29

Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	ABT*	18:19:57
Air Blank	0.000	18:20:25

\*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
02/06/2012

18:20:51

A

Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:21:52
Diagnostic Test	Pass	18:22:27
Air Blank	0.000	18:22:55
0.100 Cal Check	0.100	18:23:16
Air Blank	0.000	18:23:45
Subject Test	0.440*	18:26:52
Air Blank	0.000	18:27:26
Air Blank	0.000	18:27:53
Subject Test	XXX**	18:28:15
Air Blank	0.000	18:28:49

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GLENNHULT AGENCY GOOD YEAR POLICE DEPT,  
DATE 2/6/2012 TIME 1829  
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD AVE

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
02/06/2012

18:29:04

-----  
Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	18:29:35
0.100 Cal Check	0.099	18:29:56
Air Blank	0.000	18:30:25

Preventative Maintenance  
Performed -- Timer Reset



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. V. ELMONT AGENCY GOODYEAR PD  
DATE 2/29/2012 TIME 1743  
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD AVE.

- (A) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (A) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (A) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (A) 4. Air blank completed.
- (A) 5. Calibration check completed. Test results 0. 100 AC.
- (A) 6. Air blank completed.
- (A) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

INTOXILYZER 8000  
Location: 1111 S LITCHFIELD RD  
Serial Number: 80-001858  
Core Version: 8105.48  
02/29/2012

17:43:47

-----  
Standard Lot#: 01011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	17:44:20
0.100 Cal Check	0.100	17:44:41
Air Blank	0.000	17:45:10

Preventative Maintenance  
Performed -- Timer Reset