

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K. Miyazato 1139 AGENCY GOODYEAR P.D.  
DATE 2/1/2020 TIME 0347  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.  
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.  
( ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
() 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE

K. Miyazato 1139

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K. MIYAZATO 1139 AGENCY GOODYEAR P.D.  
DATE 2/1/2020 TIME 0026  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
() 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
() 4. Air blank completed.  
() 5. Calibration check completed. Test results 0. 097 AC.  
() 6. Air blank completed.  
() 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K. MIYAZATO 1139

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/01/2020 00:26:51

-----  
Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100  
-----

DAS: K. MIYAZATO #1139  
GOODYEAR PD  
-----

Operator: K. MIYAZATO #1139  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	00:27:37
0.100 Cal Check	0.097	00:27:58
Air Blank	0.000	00:28:28

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/01/2020

04:10:11

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

OAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Subject: SCAF, TEST  
DOB: 04/02/1973  
Sex: M Weight: 209

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:11:12
Diagnostic Test	Pass	04:11:50
Air Blank	0.000	04:12:17
0.100 Cal Check	0.097	04:12:38
Air Blank	0.000	04:13:08
Subject Test	0.263*	04:16:15
Air Blank	0.000	04:16:48
Air Blank	0.000	04:17:15
Subject Test	XXX**	04:17:28
Air Blank	0.000	04:18:02

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/01/2020

04:10:11

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

OAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:21:24
0.100 Cal Check	0.097	04:20:45
Air Blank	0.000	04:21:15

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-00031  
Core Version: 8105.48  
02/01/2020 03:52:50

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Subject: SOAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 209

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	03:53:40
Air Blank	0.000	03:54:08

\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/01/2020 03:54:27

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Subject: SOAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 209

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	03:55:39
Air Blank	0.000	03:56:08

\*RFI Detect

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/01/2020 03:56:27

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Subject: SOAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 209

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	03:57:34
Diagnostic Test	Pass	03:58:11
Air Blank	0.000	03:58:39
0.100 Cal Check	0.096	03:58:59
Air Blank	0.000	03:59:29
Subject Test	0.329*	04:02:37
Air Blank	0.000	04:03:19
Air Blank	0.000	04:03:47
Subject Test	RGE**	04:04:17
Air Blank	0.000	04:05:09

\*Deficient Sample  
\*\*Range Exceeded

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/01/2020 03:47:58

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Subject: SOAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 209

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	03:49:26
Diagnostic Test	Pass	03:50:03
Air Blank	0.000	03:50:30
0.100 Cal Check	0.097	03:50:51
Air Blank	0.000	03:51:21
Subject Test	0.000	03:51:38
Air Blank	ABT*	03:51:56
Air Blank	0.000	03:52:25

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K Miyazaki 1139 AGENCY GOODYEAR P.D.  
DATE 2/29/20 TIME 15:53  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 097 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1139

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
02/29/2020

15:53:45

Standard Lot#: 10818100A1  
Last Changed By: W. RYAN #1100

DAS: K. MIYAZATO #1139  
GOODYEAR PD

Operator: K. MIYAZATO #1139  
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	15:54:48
0.100 Cal Check	0.097	15:55:09
Air Blank	0.000	15:55:39

Preventative Maintenance  
Performed -- Timer Reset