

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. WILSON AGENCY GOODYEAR PD  
DATE 11/24/13 TIME 0054  
INTOXILYZER SERIAL # 80-001858 LOCATION 10701 SE STELLA

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 10<sup>0</sup> AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 102 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 10701 S ESTRELLA  
Serial Number: 80-001858  
Core Version: 8105.48  
01/24/2013

00:54:38



-----  
Standard Lot#: 06011100A1  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	00:55:08
0.100 Cal Check	0.102	00:55:29
Air Blank	0.000	00:55:58

Preventative Maintenance  
Performed -- Timer Reset