

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W.T. RYAN 1100 AGENCY GOODYEAR P.D.  
DATE 1-2-2014 TIME 1304  
INTOXILYZER SERIAL # 80-001858 LOCATION STATION # 2

- (V) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.10 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(S) 4. Air blank completed.  
(S) 5. Calibration check completed. Test results 0.098 AC.  
(S) 6. Air blank completed.  
(S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE W.T. Ryan 1100

INTOXILYZER 8000

Location: 10701 S ESTRELLA PKW

Serial Number: 80-001858

Core Version: 8105.48

01/02/2014

13:04:08

-----  
Standard Lot#: 32012100A1

Last Changed By: W. WILL #974

-----  
QAS: W. RYAN #1100

GOODYEAR PD

-----  
Operator: W. RYAN #1100

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      13:04:40  
0.100 Cal Check                0.098                      13:05:01  
Air Blank                      0.000                      13:05:30

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. Schubert AGENCY Goodyear PD  
DATE 1/29/2014 TIME 2246  
INTOXILYZER SERIAL # 80-001858 LOCATION 10701 S ESTEREO PKWY

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 097 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1144

INTOXILYZER 8000  
Location: 10701 S ESTRELLA PKW  
Serial Number: 80-001858  
Core Version: 8105.48  
01/29/2014

*W*

22:46:06

-----  
Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	22:46:35
0.100 Cal Check	0.097	22:46:56
Air Blank	0.000	22:47:26

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUBERBOUT AGENCY GOODYEAR PD  
DATE 11-29-2014 TIME 2247  
INTOXILYZER SERIAL # 80-001858 LOCATION 10701 S. ESTACADA HWY.

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.  
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.  
() 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
() 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

mouthwash  
purge fail - alcohol close to tube -  
removed after restarting good.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 10701 S ESTRELLA PKW  
Serial Number: 80-001858  
Core Version: 8105.48  
01/29/2014

22:47:59

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	22:48:24
Air Blank	IPS**	22:48:37

\*RFI Detect

\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 10701 S ESTRELLA PKW  
Serial Number: 80-001858  
Core Version: 8105.48  
01/29/2014

22:48:41

A

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:49:31
Diagnostic Test	Pass	22:50:06
Air Blank	0.000	22:50:34
0.100 Cal Check	0.098	22:50:54
Air Blank	0.000	22:51:24
Subject Test	0.000	22:51:42
Air Blank	ABT*	22:51:49
Air Blank	0.000	22:52:17

\*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 10701 S ESTRELLA PKW  
Serial Number: 80-001858  
Core Version: 8105.48  
01/29/2014

22:52:23

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:53:14
Diagnostic Test	Pass	22:53:49
Air Blank	0.000	22:54:17
0.100 Cal Check	0.098	22:54:37
Air Blank	0.000	22:55:07
Subject Test	0.126*	22:58:14
Air Blank	PUR**	22:59:01
Air Blank	PUR**	22:59:48

\*Deficient Sample

\*\*Purge Fail

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 10701 S ESTRELLA PKW  
Serial Number: 80-001858  
Core Version: 8105.48  
01/29/2014

22:59:54

Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Subject: TEST,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:00:42
Diagnostic Test	Pass	23:01:17
Air Blank	0.000	23:01:44
0.100 Cal Check	0.098	23:02:05
Air Blank	0.000	23:02:35
Subject Test	XXX*	23:02:57
Air Blank	0.000	23:03:30

\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUBBART AGENCY GOODYEAR PD  
DATE 1/29/2014 TIME 2303  
INTOXILYZER SERIAL # 80-001858 LOCATION 10701 S ESTERNA HWY

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 097 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: 10701 S ESTRELLA PKW  
Serial Number: 80-001858  
Core Version: 8105.48  
01/29/2014

23:03:51

-----  
Standard Lot#: 32012100A1  
Last Changed By: W. WILL #974  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	23:04:20
0.100 Cal Check	0.097	23:04:41
Air Blank	0.000	23:05:10

Preventative Maintenance  
Performed -- Timer Reset