



ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER SERVICE REQUEST
SEND ALL COPIES TO DPS CRIME LABORATORY

QUALITY ASSURANCE SPECIALIST

DATE 7/6/07	SERIAL NUMBER (List simulator S / N also) INTOX: 80-00 1858 SIMULATOR:	FIELD LOCATION Good year PD
DPS EQUIPMENT NUMBER	SUBMITTING AGENCY'S FULL NAME (NO INITIALS)	
QUALITY ASSURANCE SPECIALIST BILL WELLS		PHONE NUMBER ()
QUALITY ASSURANCE SPECIALIST MAILING ADDRESS		CITY STATE ZIP CODE

Was a S.Q.A.P. completed prior to removal from service?

☐ No, Explain: _____

☐ Yes, Date: _____ Date out of service: _____

☒ No ☐ Yes Do records need to be downloaded (ACTIVE C.O.B.R.A.. ONLY)

EQUIPMENT SENT:

☒ INTOXILYZER ☒ POWER CORD ☒ BREATH HOSE ☐ EXTERNAL PRINTER ☐ SIMULATOR ☒ KEYBOARD ☐ MODEM
☐ HEATED SIMULATOR TUBE ☐ BRACKET ☒ OTHER **GAS**

PROBLEM DESCRIPTION: (What is the instrument doing or not doing? Include observations such as processor errors or other displayed messages.)

Temp Ren Error

LAB USE ONLY

CORRECTIVE ACTION

Replace BREATH HOSE
verify CALIBRATION & FUNCTION CHECK

COMMENTS

TECHNICIAN'S NAME

BADGE / ID NO. TITLE

DATE

CALLED FOR PICKUP

DATE

CONTACT PERSON

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number : 80-001858

07/13/2007

09:12:32

Standard Lot#: 536401I

Last Changed By: M. SLONEKER #6559

QAS: J. ROGERS #1123

GOODYEAR PD

Operator: J. TRAINING #1111

AZ DPS

Subject: FOUR,SAMPLE,T

DOB: 03/01/1965

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:13:15
Diagnostic Test	Pass	09:13:42
Air Blank	0.000	09:14:10
0.100 Cal Check	0.099	09:14:28
Air Blank	0.000	09:14:58
Subject Test	0.188	09:15:30
Air Blank	0.000	09:16:03
Five Minute Wait		
Air Blank	0.000	09:20:57
Subject Test	0.188	09:21:20
Air Blank	0.000	09:21:52
0.100 Cal Check	0.099	09:22:11
Air Blank	0.000	09:22:40
Diagnostic Test	Pass	09:23:08

Successfully Completed Test Sequence