

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

07/03/2009

20:22:03


Standard Lot#: 562576

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	20:22:30
Cal Check	0.100	20:22:49
Air Blank	0.000	20:23:19



Operator's Signature

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

07/03/2009

20:18:33

Standard Lot#: 808401G

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	20:19:00
Cal Check	0.100	20:19:19
Air Blank	0.000	20:19:48



Operator's Signature

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GULBAVET AGENCY GOODYEAR PD
DATE 7/18/09 TIME 1909
INTOXILYZER SERIAL # 80-001858 LOCATION 6PD

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 103 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

4. Air blank completed.

5. Calibration check completed. Test results 0. 103 AC.

6. Air blank completed.

7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

07/18/2009

19:09:11


Standard Lot#: 562576

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	19:09:38
Cal Check	0.103	19:09:57
Air Blank	0.000	19:10:27



Operator's Signature

1144

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GURMAN AGENCY GOODYEAR PD
DATE 7/28/09 TIME 1313
INTOXILYZER SERIAL # 80-001858 LOCATION GPD

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

07/28/2009

13:12:42


Standard Lot#: 562576

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	13:13:09
Cal Check	0.101	13:13:28
Air Blank	0.000	13:13:58



Operator's Signature