

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. LEWIS/1144 AGENCY GOODRICH AS  
DATE 3/14/2013 TIME 1701  
INTOXILYZER SERIAL # PD-001858 LOCATION 10701 S ESTRELLA PKWY

- (P) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
4. Air blank completed.  
5. Calibration check completed. Test results 0. 100 AC.  
6. Air blank completed.  
(T) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE M. Lewis

INTOXILYZER 8000  
Location: 10701 S ESTRELLA  
Serial Number: 80-001858  
Core Version: 8105.48  
03/14/2013

17:01:47

-----  
Standard Lot#: 13712100A2  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	17:02:37
0.100 Cal Check	0.100	17:02:58
Air Blank	0.000	17:03:28

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000  
Location: 1111 S LICHFIELD RD  
Serial Number: 80-000312  
Core Version: 8105.48  
03/14/2013

17:49:40

-----  
Standard Lot#: 13712100A2  
Last Changed By: M. GUILBAULT #1144

-----  
QAS: M. GUILBAULT #1144  
GOODYEAR PD

-----  
Operator: M. GUILBAULT #1144  
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	17:50:10
0.100 Cal Check	0.094	17:50:31
Air Blank	0.000	17:51:01

Preventative Maintenance  
Performed -- Timer Reset