

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST MICHAEL GUDENOUR #1144 AGENCY GOODYEAR PD

DATE 5/22/09 TIME 0825

INTOXILYZER SERIAL # 80-001858 LOCATION 175N 145TH AVE, GOODYEAR

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 100 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: GOODYEAR POLICE DEPT  
Serial Number: 80-001858  
05/22/2009

08:25:18

-----  
Standard Lot#: 808401G  
Last Changed By: M. GUILBAULT #1144

-----  
QAS: M. GUILBAULT #1144  
GOODYEAR PD

-----

Test	g/210L	Time
Air Blank	0.000	08:25:45
Cal Check	0.100	08:26:04
Air Blank	0.000	08:26:33


  
-----  
Operator's Signature

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST WILL AGENCY Grosvonts PD  
DATE 5/25/2009 TIME 15:12  
INTOXILYZER SERIAL # 80-1858 LOCATION Grosvonts PD

- (W) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (W) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(W) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(W) 4. Air blank completed.  
(W) 5. Calibration check completed. Test results 0. .099 AC.  
(W) 6. Air blank completed.  
(W) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

05/25/2009

15:11:36

-----  
Standard Lot#: 808401G

Last Changed By: M. GUILBAULT #1144

-----  
OAS: W. WIEL #974

GOODYEAR PD

-----  
Test                      5/210L                      Time  
-----  
Air Blank                      0.000                      15:12:03  
Cal Check                      0.099                      15:12:22  
Air Blank                      0.000                      15:12:51

-----  
Operator's Signature

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPT

Serial Number: 80-001858

05/25/2009

15:15:06

-----  
Standard Lot#: 808401G

Last Changed By: M. GUILBAULT #1144

-----  
QAS: M. GUILBAULT #1144

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	15:15:34
Cal Check	0.100	15:15:52
Air Blank	0.000	15:16:22



\_\_\_\_\_  
Operator's Signature