

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. VELMANT AGENCY GOOD TERRA PD
DATE 5/10/2013 TIME 1809
INTOXILYZER SERIAL # 80-001858 LOCATION 10701 S ESTRELLA PKWY

- (P) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (P) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (P) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (P) 4. Air blank completed.
- (P) 5. Calibration check completed. Test results 0. 100 AC.
- (P) 6. Air blank completed.
- (P) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE A. 1144

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-001858
Core Version: 8105.48
05/10/2013

18:09:30 A

Standard Lot#: 13712100A2
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	18:10:03
0.100 Cal Check	0.100	18:10:24
Air Blank	0.000	18:10:54

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUELFUSOULT AGENCY GOODYEAR PD
DATE 5/10/13 TIME 1811
INTOXILYZER SERIAL # 80-001858 LOCATION 10701 S ESTRELLA PKWY

(~~DI~~) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (~~DI~~) 1. Clock time check.
(~~DI~~) 2. Date check.

OPERATIONAL TESTS

- (~~DI~~) 1. Alcohol-free subject test result 0. 000 AC.
(~~DI~~) 2. Error recognition logic system functioning.

Not a Successfully Completed Test Sequence printed

- (~~DI~~) 3. Proper sample recognition system.

Not a Successfully Completed Test Sequence printed

Deficient sample printed.

- (~~DI~~) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS

SIGNATURE WA. 144

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-001858
Core Version: 8105.48
05/10/2013

18:11:02

Standard Lot#: 13712100A2
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:11:47
Diagnostic Test	Pass	18:12:22
Air Blank	0.000	18:12:50
0.100 Cal Check	0.100	18:13:10
Air Blank	0.000	18:13:40
Subject Test	0.000	18:13:53
Air Blank	IPS*	18:13:59
Air Blank	RFI**	18:14:07

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-001858
Core Version: 8105.48
05/10/2013

18:14:17

Standard Lot#: 13712100A2
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 199

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:15:05
Diagnostic Test	Pass	18:15:41
Air Blank	0.000	18:16:08
0.100 Cal Check	0.099	18:16:29
Air Blank	0.000	18:16:58
Reference	ABT*	18:17:05
Air Blank	0.000	18:17:33

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-001858
Core Version: 8105.48
05/10/2013

18:17:42

Standard Lot#: 13712100A2
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:18:36
Diagnostic Test	Pass	18:19:11
Air Blank	0.000	18:19:38
0.100 Cal Check	0.100	18:19:59
Air Blank	0.000	18:20:29
Subject Test	0.008*	18:23:36
Air Blank	0.000	18:24:07
Air Blank	0.000	18:24:35
Subject Test	XXX**	18:24:57
Air Blank	0.000	18:25:31

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. VELLEBAULT AGENCY GOODYEAR PD
DATE 5/10/2013 TIME 1826
INTOXILYZER SERIAL # FD 001858 LOCATION 10701 SESTRINELLA PKWY

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: 10701 S ESTRELLA
Serial Number: 80-001858
Core Version: 8105.48
05/10/2013

18:26:00

Standard Lot#: 13712100A2
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	18:26:29
0.100 Cal Check	0.099	18:26:50
Air Blank	0.000	18:27:19

Preventative Maintenance
Performed -- Timer Reset