

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Rogers AGENCY GOODYEAR P.D.
DATE 10-11-07 TIME 1851
INTOXILYZER SERIAL # 80-001858 LOCATION GOODYEAR P.D.

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] # 1123

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPA

Serial Number: 80-001858

10/11/2007

18:51:46


Standard Lot#: 536401I

Last Changed By: M. SLONEKER #6559

QAS: J. ROGERS #1123

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	18:52:13
Cal Check	0.100	18:52:32
Air Blank	0.000	18:53:02

 #1123

Operator's Signature

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPA

Serial Number: 80-001858

10/25/2007

11:07:38


Standard Lot#: 706502I

Last Changed By: W. WILL #974

QAS: W. WILL #974

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	11:08:05
Cal Check	0.100	11:08:24
Air Blank	0.000	11:08:53



Operator's Signature

STANDARD CHANGE
AFTER

80-001858

10/25/2007

**Scott Specialty Gases, Inc.****CERTIFICATE OF ANALYSIS****EBS™ - Ethanol Breath Standard**

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order: 370933
Scott Project # 0199359
Lot No.: 7065021
Expiration Date:
March 9, 2009

<u>REFERENCE STANDARD</u>	<u>CYLINDER #</u>	<u>EXPIRATION DATE</u>	<u>CONCENTRATION PPM</u>
N.I.S.T. Traceable Standards*	ALM049986	05/18/2007	262.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE**CONCENTRATION**COMPONENTS

Ethanol
Nitrogen

ppm (BrAC)

260.5 (0.100)
Balance

ACCURACY

+/- 2%

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst:

Date: 03/14/2007

INTOXILYZER 8000

Location: GOODYEAR POLICE DEPA

Serial Number: 80-001858

10/25/2007

11:03:46


Standard Lot#: 536401I

Last Changed By: M. SLONEKER #6559

QAS: W. WILL #974

GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	11:04:14
Cal Check	0.100	11:04:32
Air Blank	0.000	11:05:02



Operator's Signature

STANDARD CHANGE
PRIOR