

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)


ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. G. GLEIBERT AGENCY GOODYEAR PD  
DATE 10/6/13 TIME 0837  
INTOXILYZER SERIAL # 80-001858 LOCATION GPD #2

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_ 

INTOXILYZER 8000  
Location: GOODYEAR PD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/06/2010

08:37:07

-----  
Standard Lot#: 767153  
Last Changed By: M. GUILBAULT #1144  
-----

QAS: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Operator: M. GUILBAULT #1144  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	08:37:53
0.100 Cal Check	0.098	08:38:14
Air Blank	0.000	08:38:44

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Will AGENCY Goodyear PD  
DATE 10/10/2010 TIME 1726  
INTOXILYZER SERIAL # PO-001858 LOCATION Goodyear PD

- (W) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (W) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(W) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(W) 4. Air blank completed.  
(W) 5. Calibration check completed. Test results 0. 1092 AC.  
(W) 6. Air blank completed.  
(W) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

DPS Form Exh G-2 (Rev 05-01)

\* O-RING SEAL FOUND TO BE BROKEN



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUILMANT 1144 AGENCY GOODYEAR PD  
DATE 10/26/2010 TIME 1410  
INTOXILYZER SERIAL # 80-001858 LOCATION 6PA # 2

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution of 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 1144

INTOXILYZER 8000  
Instrument Initialization  
13:37:56 10/26/2010

*M. Guilbault*

INTOXILYZER 8000  
Location: GOODYEAR PD  
Serial Number: 80-001858  
Core Version: 8105.48  
10/26/2010

14:10:13

Standard Lot#: 767153  
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144  
GOODYEAR PD

Operator: M. GUILBAULT #1144  
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	14:10:49
0.100 Cal Check	0.101	14:11:10
Air Blank	0.000	14:11:40

Preventative Maintenance  
Performed -- Timer Reset