

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUILBAULT AGENCY GOODHEAR PD
DATE 9/27/2011 TIME 1753
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 100 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. GUELBAULT AGENCY GOODYEAR PD
DATE 9/27/2011 TIME 1756
INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD LN

(~~1~~) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- (~~1~~) 1. Clock time check.
(~~1~~) 2. Date check.

OPERATIONAL TESTS

- (~~1~~) 1. Alcohol-free subject test result 0. 000 AC.
(~~1~~) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(~~1~~) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(~~1~~) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES X NO _____

COMMENTS

SIGNATURE  1144

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
09/27/2011

17:56:23

[Handwritten signature] 1144

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST

DOB: 01/01/1990

Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:57:15
Diagnostic Test	Pass	17:57:50
Air Blank	0.000	17:58:17
0.100 Cal Check	0.099	17:58:38
Air Blank	0.000	17:59:08
Subject Test	0.000	17:59:24
Air Blank	RFI*	17:59:33
Air Blank	IPS**	17:59:42

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: 1111 S LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
09/27/2011

18:00:13

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Subject: TEST,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:01:07
Diagnostic Test	Pass	18:01:42
Air Blank	ABT*	18:01:50
Air Blank	0.000	18:02:18

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

09/27/2011

18:02:30

Standard Lot#: 01011100A1

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Subject: TEST,TEST

DOB: 01/01/1990

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:03:39
Diagnostic Test	Pass	18:04:14
Air Blank	0.000	18:04:42
0.100 Cal Check	0.099	18:05:03
Air Blank	0.000	18:05:32
Subject Test	0.048*	18:08:40
Air Blank	0.000	18:09:12
Air Blank	0.000	18:09:39
Subject Test	XXX**	18:09:58
Air Blank	0.000	18:10:32

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M GUILBAULT AGENCY GOODYEAR PD

DATE 9/27/2011 TIME 1811

INTOXILYZER SERIAL # 80-001858 LOCATION 1111 S LITCHFIELD RD

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 100 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000

Location: 1111 S LITCHFIELD RD

Serial Number: 80-001858

Core Version: 8105.48

09/27/2011

18:11:47

Standard Lot#: 01011100A1

Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144

GOODYEAR PD

Operator: M. GUILBAULT #1144

GOODYEAR PD

Test g/Z10L Time

Air Blank 0.000 18:12:17
0.100 Cal Check 0.100 18:12:38
Air Blank 0.000 18:13:07

Preventative Maintenance
Performed -- Timer Reset