

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear  
DATE 4/27/2022 TIME 0354  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145 TH AVE

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE K Miyazato 1139

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

04/27/2022

03:54:51

---

Standard Lot#: 27220100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	03:55:48
Diagnostic Test	Pass	03:56:23
Air Blank	0.000	03:56:51
0.100 Cal Check	0.099	03:57:11
Air Blank	0.000	03:57:41
Subject Test	0.000	03:57:55
Air Blank	ABT*	03:58:04
Air Blank	0.000	03:58:32

\*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
04/27/2022 03:58:36

---

Standard Lot#: 27220100A1  
Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139  
GOODYEAR PD

---

Operator: K. MIYAZATO #1139  
GOODYEAR PD

---

Subject: SQAP, TEST  
DOB: 04/02/1973  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	03:59:12
Air Blank	0.000	03:59:41

\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

04/27/2022

03:59:44

---

Standard Lot#: 27220100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	04:00:20
Air Blank	0.000	04:00:49

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
04/27/2022

04:00:55

---

Standard Lot#: 27220100A1  
Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139  
GOODYEAR PD

---

Operator: K. MIYAZATO #1139  
GOODYEAR PD

---

Subject: SQAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:01:47
Diagnostic Test	Pass	04:02:22
Air Blank	0.000	04:02:49
0.100 Cal Check	0.099	04:03:10
Air Blank	0.000	04:03:40
Subject Test	0.070*	04:06:47
Air Blank	0.000	04:07:19
Air Blank	0.000	04:07:46
Subject Test	XXX**	04:07:59
Air Blank	0.000	04:08:32

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goupy SR  
DATE 4/27/2022 TIME \_\_\_\_\_  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145TH AVE

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
() 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
() 4. Air blank completed.  
() 5. Calibration check completed. Test results 0.098 AC.  
() 6. Air blank completed.  
() 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_ 1139

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
04/27/2022 04:08:49

-----  
Standard Lot#: 27220100A1  
Last Changed By: K. MIYAZATO #1139  
-----

QAS: K. MIYAZATO #1139  
GOODYEAR PD  
-----

Operator: K. MIYAZATO #1139  
GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	04:09:23
0.100 Cal Check	0.098	04:09:43
Air Blank	0.000	04:10:13

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

04/27/2022

04:13:45

-----  
Standard Lot#: 28021100A1

Last Changed By: K. MIYAZATO #1139

-----  
QAS: K. MIYAZATO #1139

GOODYEAR PD

-----  
Operator: K. MIYAZATO #1139

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      04:14:16  
0.100 Cal Check                      0.100                      04:14:37  
Air Blank                      0.000                      04:15:07

Preventative Maintenance  
Performed -- Timer Reset





7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

## Certificate of Analysis

**Certificate ID:** 14350  
**Part #:** BAC105L100T  
**Cylinder Size:** 105L  
**Lot Number:** 28021100A1  
**Expiration:** 12/5/2023

**0.100 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)


<b>Component:</b>	<b>Reported Concentration:</b>	<b>Analytical Accuracy (U, k=2):</b>	<b>Analytical Method:</b>	<b>Distributed by:</b>
Ethanol	260 ppm	+/-0.002 BAC(G/210L)	NDIR	<b>CMI Inc.</b> 316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 <a href="http://www.alcoholtest.com">www.alcoholtest.com</a>
Nitrogen	Balance	[5.2 ppm]		

\*Traceable to:  
Certified Reference Material - 262.4 µmol/mol  
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

11-09-2021  
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

	Chemical Substance	Chemical Family	Trade Names
Ethanol	ETHYL ALCOHOL, 100%	hydroxyls, aliphatic, alcohols, aliphatic	ETHANOL; ETHYL ALCOHOL; 4; ALCOHOL; ALCOHOL ANHYDROUS; ALGRAIN; ANHYDROL; Absolute alcohol; Anhydrous ethanol; Ethanol denatured; Fermentation alcohol; Grain alcohol; 1-Hydroxyethane; Methyl carbinol; Ethyl alcohol anhydrous; Absolute ethanol; Denatured ethanol; ETHYL HYDRATE; ETHYL HYDROXIDE; JAYSOL; TEGSOL; STCC 4909159; UN 1170; C2H6O
Nitrogen	NITROGEN, COMPRESSED GAS	Inorganic gases	DIATOMIC NITROGEN; DINITROGEN; NITROGEN; NITROGEN-14; NITROGEN GAS; UN 1066; N2

### Section 1: Product and Company Identification

ILMO Products Company  
7 Eastgate Drive, Jacksonville, Illinois 62650  
217-245-2183 800-424-9300 (Chemtrek)

Fax 217-243-7634  
E-mail: info@ilmoproducts.com  
Web: www.ilmoproducts.com

Product Code: Ethanol in Nitrogen  
Part Number: BAC

### Section 2: Hazards Identification



Hazard Classification:  
Gases Under Pressure

Hazard Statements:  
Contains gas under pressure; may explode if heated

Precautionary Statements

Storage:  
Protect from sunlight.  
Store in well-ventilated place.

### Section 3: Composition/Information on Ingredients

	CAS #	Concentration
Ethanol	64-17-5	5-500 ppm
Nitrogen	7727-37-9	Balance

ILMO Products Company  
Generated by the SDS Manager from AsteRisk, LLC. All Rights Reserved

page 1 of 6  
Generated: 02/08/2017 10:18:45

### Section 4: First Aid Measures

	Skin Contact	Eye Contact	Ingestion	Inhalation	Note to Physicians
Ethanol	Wash skin with soap and water for at least 15 minutes while removing contaminated clothing and shoes. Get medical attention, if needed. Thoroughly clean and dry contaminated clothing and shoes before reuse.	Flush eyes with plenty of water for at least 15 minutes. Then get immediate medical attention.	Contact local poison control center or physician immediately. Never make an unconscious person vomit or drink fluids. When vomiting occurs, keep head lower than hips to help prevent aspiration. If person is unconscious, turn head to side. Get medical attention immediately.	If adverse effects occur, remove to uncontaminated area. Give artificial respiration if not breathing. Get immediate medical attention.	For ingestion, consider gastric lavage.
Nitrogen	Wash exposed skin with soap and water.	Flush eyes with plenty of water.	If a large amount is swallowed, get medical attention.	If adverse effects occur, remove to uncontaminated area. Give artificial respiration if not breathing, if breathing is difficult, oxygen should be administered by qualified personnel. Get immediate medical attention.	For inhalation, consider oxygen.

### Section 5: Fire Fighting Measures

	Suitable Extinguishing Media	Products of Combustion	Protection of Firefighters
Ethanol	Alcohol resistant foam, carbon dioxide, regular dry chemical, water, alcohol resistant foam. Large fires: Use alcohol-resistant foam or food with fine water spray.	Carbon monoxide, carbon dioxide, and toxic and irritating fumes	<ul style="list-style-type: none"> <li>Any supplied-air respirator with full facepiece and operated in a pressure-demand or other positive-pressure mode in combination with a separate escape supply.</li> <li>Any supplied-air respirator with full facepiece and operated in a pressure-demand or other positive-pressure mode in combination with a separate escape supply.</li> </ul>
Nitrogen	Non-flammable. Use suitable extinguishing media for surrounding fire. Cylinders may rupture or explode if exposed to heat.	Non-flammable	Respiratory protection may be needed for frequent or heavy exposure.

### Section 6: Accidental Release Measures

	Personal Precautions	Environmental Precautions	Methods for Containment
Ethanol	Keep unnecessary people away, isolate hazard area and deny entry. Stay upwind and keep out of low areas.	Avoid heat, flames, sparks and other sources of ignition.	Stop leak if possible without personal risk. Reduce vapors with water spray. Remove sources of ignition.
Nitrogen	Keep unnecessary people away, isolate hazard area and deny entry. Stay upwind and keep out of low areas.	No significant effects from contamination expected.	Stop leak if possible without personal risk.

ILMO Products Company  
Generated by the SDS Manager from AsteRisk, LLC. All Rights Reserved

page 2 of 6  
Generated: 02/08/2017 10:18:45

	Methods for Cleanup	Other Information
Ethanol	Small spills: Absorb with sand or other noncombustible material. Collect spilled material in appropriate container for disposal. Large spills: Dike for later disposal.	Not available
Nitrogen	N/A	N/A

### Section 7: Handling and Storage

	Handling	Storage
Ethanol	Store and use with adequate ventilation. Firmly secure cylinders upright to keep them from falling or being knocked over. Screw valve protection cap firmly in place by hand. Store only where temperature will not exceed 125°F (52°C). Store full and empty cylinders separately. Use a first-in, first-out inventory system to prevent storing full cylinders for long periods.	Do not get liquid in eyes, on skin, or clothing. Protect cylinders from damage. Use a suitable hand truck to move cylinders; do not drag, roll, slide, or drop. Open valve slowly. Close cylinder valve after each use, keep closed even when empty. If valve is hard to open, discontinue use and contact your supplier.
Nitrogen	Store and handle in accordance with all current regulations and standards. Subject to storage regulations: U.S. OSHA 29 CFR 1910.101.	Keep separated from incompatible substances.

### Section 8: Exposure Controls/Personal Protection

	Exposure Guidelines
Ethanol	ETHYL ALCOHOL, 100%: ETHYL ALCOHOL (ETHANOL): 1000 ppm (1900 mg/m <sup>3</sup> ) OSHA TWA 1000 ppm ACGIH TWA 1000 ppm (1900 mg/m <sup>3</sup> ) NIOSH recommended TWA 10 hour(s)
Nitrogen	NITROGEN, COMPRESSED GAS: NITROGEN: ACGIH (simple asphyxiant)

Engineering Controls  
Handle only in fully enclosed systems.

	Eye Protection	Skin Protection	Respiratory Protection
Ethanol	Wear splash resistant safety goggles. Provide an emergency eye wash fountain and quick drench shower in the immediate work area.	Wear appropriate chemical resistant clothing.	Any supplied-air respirator with full facepiece and operated in a pressure-demand or other positive-pressure mode in combination with a separate escape supply.
Nitrogen	Eye protection not required, but recommended.	Protective clothing is not required.	Respiratory protection may be needed for frequent or heavy exposure.

General Hygiene considerations

- Avoid breathing vapor or mist
- Avoid contact with eyes and skin
- Wash thoroughly after handling and before eating or drinking

### Section 9: Physical and Chemical Properties

	Physical State	Appearance	Color	Change in Appearance	Physical Form	Odor	Taste
Ethanol	Liquid	Clear	Colorless	N/A	Volatile liquid	Pleasant odor	Burning taste
Nitrogen	Gas	Clear	Colorless	N/A	Gas	Odorless	Tasteless

	Flash Point	Flammability	Partition Coefficient	Autoignition Temperature	Upper Explosive Limits	Lower Explosive Limits
Ethanol (CC)	55 F (13 C)	IB	Not available	685 F (363 C)	0.19	0.033
Nitrogen	Not flammable	Not available	Not available	Nonflammable	Nonflammable	Nonflammable

	Boiling Point	Freezing Point	Vapor Pressure	Vapor Density	Specific Gravity	Water Solubility	pH	Odor Threshold	Evaporation Rate	Viscosity
Ethanol	172 F (78 C)	-179 F (-117 C)	40 mmHg @ 19 C	1.59 (Air=1)	0.7893	Soluble	Not available	Not available	5-10 ppm	1.17 mPa.s (1.17 centipoises) @ 20 C, 1.074 mPa.s (1.074 centipoises) @ 25 C, 0.01787 cP @ 27 C
Nitrogen	-321 F (-196 C)	-346 F (-210 C)	760 mmHg @ -196 C	0.967 (Air=1)	Not applicable	1.6% @ 20 C	Not applicable	Not available	Not applicable	0.01787 cP @ 27 C

ILMO Products Company  
Generated by the SDS Manager from AsteRisk, LLC. All Rights Reserved

page 3 of 6  
Generated: 02/08/2017 10:18:45

### Section 10: Stability and Reactivity

	Stability	Conditions to Avoid	Incompatible Materials
Ethanol	Stable at normal temperatures and pressure.	Stable at normal temperatures and pressure.	Halo carbons, metals, metal salts, oxidizing materials, halogens, peroxides, acids, metal oxides, bases, combustible materials
Nitrogen	Stable at normal temperatures and pressure.	Stable at normal temperatures and pressure.	Metals, oxidizing materials

	Hazardous Decomposition Products	Possibility of Hazardous Reactions
Ethanol	Oxides of carbon	Will not polymerize.
Nitrogen	Oxides of nitrogen	Will not polymerize.

### Section 11: Toxicology Information

	Acute Effects	Inhalation
Ethanol	Oral LD50: 7 gm/kg oral-rat LD50	LD50 (dermal, rabbit): greater than 15800 mg/kg (cited as greater than 20 mL/kg); at 20 mL/kg, 1/4 rabbits died
Nitrogen	Not available	Not available

	Eye Irritation	Skin Irritation	Sensitization
Ethanol	Irritation, tearing	Mild irritation, rash	Respiratory tract irritation, skin irritation, eye irritation, liver damage, central nervous system depression
Nitrogen	Contact with rapidly expanding gas may cause burns or frostbite	No information on significant adverse effects	Difficulty breathing

	Chronic Effects	Mutagenicity	Reproductive Effects	Developmental Effects
Ethanol	Carcinogenicity: NTP: Known Human Carcinogen (Alcoholic beverages); IARC: Human Sufficient Evidence, Group 1 (Alcoholic beverages); Animal Inadequate Evidence: ACGIH: A4 - Not Classifiable as a Human Carcinogen	Available.	Available.	No data
Nitrogen	Not hazardous	Not available	Not available	No data

ILMO Products Company  
Generated by the SDS Manager from AsteRisk, LLC. All Rights Reserved

page 4 of 6  
Generated: 02/08/2017 10:18:45