

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato AGENCY Goodyear PD  
DATE 9/20/22 TIME 2109  
INTOXILYZER SERIAL # 90-0003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration -100 AC

DIAGNOSTIC TESTS

1. Clock time check  
 2. Date check


OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
5. Standard Calibration Check:  
0.100 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS: GAS COMPLETED

SIGNATURE 

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37950920220005

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 09/19/2022  
Last Annual Maintenance: 08/03/2022

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 09/20/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 20:45  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402376471  
Expiration Date: 06/02/2024  
Bottle #: 44  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	21:09:47
Diagnostic Test	PASS	21:10:21
Air Blank	0.000	21:10:58
Calibration Chk	0.100	21:11:20
Air Blank	0.000	21:12:00
Subject Test 1	DEF*	21:15:14
Air Blank	0.000	21:15:56
Air Blank	0.000	21:16:36
Subject Test 2	0.000	21:17:00
Air Blank	0.000	21:17:41
Wait		21:21:45
Air Blank	0.000	21:22:22
Subject Test 3	INV**	21:22:44
Air Blank	RFI***	21:23:08

## RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

QAS COMPLETED

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

INVOICE # : 1109071819

CMI  
316 EAST NINTH STREET  
OWENSBORO, KY 42303

PO # : 8003800

CMI ITEM # : 340287

Date: May 24, 2022

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

PRODUCT EXPIRATION: June 02, 2024

CALGAZ LOT#: 302-402376471

ETHANOL IN NITROGEN

COMPONENT	PPM	( BrAC )
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38434	260.4

\*Certification traceability is recognized by NIST (USA) through the CIPM (International Committee of Weights and Measures) MRA (Mutual Recognition Arrangement)

Manufactured Date: March 02, 2022

CALGAZ CYLINDER SIZE: 6DM

APPROVED BY: \_\_\_\_\_



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**  
821 Chesapeake Drive, Cambridge, MD 21613-0149  
Phone: (410) 228-6400 Fax: (410) 228-4251

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear PD  
DATE 9/30/22 TIME 1523  
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check  
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
5. Standard Calibration Check:  
0.098 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGNATURE K Miyazato 1139

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37950930220012

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 09/20/2022  
Last Annual Maintenance: 08/03/2022

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 09/30/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 05:00  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402376471  
Expiration Date: 06/02/2024  
Bottle #: 44  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	05:23:01
Diagnostic Test	PASS	05:23:37
Air Blank	0.000	05:24:16
Calibration Chk	0.097	05:24:37
Air Blank	0.000	05:25:17
Subject Test 1	DEF*	05:28:29
Air Blank	0.000	05:29:11
Air Blank	0.000	05:29:51
Subject Test 2	0.000	05:30:13
Air Blank	0.000	05:30:57
Wait		05:35:01
Air Blank	0.000	05:35:38
Subject Test 3	INV**	05:36:01
Air Blank	RFI***	05:36:32

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 09/30/2022

Test Time: 05:37

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 302-402376471

Standard Value: 0.100

Expiration Date: 06/02/2024

Bottle #: 44

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	05:37:43
Calibration Chk	0.098	05:38:05
Air Blank	0.000	05:38:45

No RFI Detected