

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa AGENCY Goodyear PD  
DATE 12/12/2022 TIME 2249  
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.00 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.099 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGNATURE K Miyazawa 1139

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-003795

Test Date: 12/12/2022

Test Time: 23:10

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 28021100A1

Standard Value: 0.100

Expiration Date: 12/05/2023

Bottle #: 001

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	23:10:35
Calibration Chk	0.099	23:10:56
Air Blank	0.000	23:11:36

No RFI Detected

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37951212220083

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 12/12/2022  
Last Annual Maintenance: 08/03/2022

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 12/12/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 22:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 28021100A1  
Expiration Date: 12/05/2023  
Bottle #: 001  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	22:49:22
Diagnostic Test	PASS	22:49:58
Air Blank	0.000	22:50:34
Calibration Chk	0.100	22:50:56
Air Blank	0.000	22:51:36
Subject Test 1	NSG*	22:54:48
Air Blank	0.000	22:55:31
Air Blank	0.000	22:56:12
Subject Test 2	0.000	22:56:33
Air Blank	0.000	22:57:14
Wait		23:01:19
Air Blank	0.000	23:01:56
Subject Test 3	INV**	23:02:41
Air Blank	RFI***	23:03:06

## RESULTS

## EXCEPTION MESSAGES

\*No Sample Given  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* No Sample Given  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37951212220082

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 10/30/2022  
Last Annual Maintenance: 08/03/2022

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 12/12/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 22:20  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 28021100A1  
Expiration Date: 12/05/2023  
Bottle #: 001  
Last Changed By: MIYAZATO, KORY

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	22:45:22
Diagnostic Test	DIA*	22:45:26
Air Blank	0.000	22:46:03

\*00004000008  
Cell Temp Out Of Range  
IRPCM Status Error

Permit #: 30215

## RESULTS

## EXCEPTION MESSAGES

\* Diagnostic Fail

## OPERATOR COMMENTS

/

No RFI Detected

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*



ARIZONA DEPARTMENT OF PUBLIC SAFETY  

---

INTOXILYZER MAINTENANCE REPORT

CR NO. I900037950002

Page 1 of 1

**AGENCY** Goodyear Police Department  
Goodyear, AZ 85338  
**OFFICER** MIYAZATO, #1139  
**DATE** November 21, 2022



**INSTRUMENT**

90003795

**INSTRUMENT LOCATION**

Goodyear Police Department

**MAINTENANCE REQUESTED**

Repair

**MAINTENANCE PERFORMED**

***Parts Replaced:***

Breath Hose

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

---

JOSEPH M SLOWINSKI, #6101, Forensic Services  
Technician  
Central Regional Crime Laboratory  
2102 W. Encanto Blvd., Phoenix, AZ 85009

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY Goodyear PD  
DATE 12/30/20 TIME 0533  
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample - Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

---

---

---

SIGNATURE K Miyazawa 1139

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37951230220102

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 12/12/2022  
Last Annual Maintenance: 08/03/2022

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 12/30/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 05:15  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215  
STANDARD INFORMATION  
Standard Value: 0.100  
Standard Lot #: 28021100A1  
Expiration Date: 12/05/2023  
Bottle #: 001  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	05:33:33
Diagnostic Test	PASS	05:34:08
Air Blank	0.000	05:34:45
Calibration Chk	0.100	05:35:07
Air Blank	0.000	05:35:46
Subject Test 1	NSG*	05:39:04
Air Blank	0.000	05:39:47
Air Blank	0.000	05:40:27
Subject Test 2	0.000	05:40:49
Air Blank	0.000	05:41:34
Wait		05:45:37
Air Blank	0.000	05:46:14
Subject Test 3	INV**	05:46:45
Air Blank	RFI***	05:47:27

## RESULTS

## EXCEPTION MESSAGES

\*No Sample Given  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* No Sample Given  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 12/30/2022

Test Time: 05:48

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 28021100A1

Standard Value: 0.100

Expiration Date: 12/05/2023

Bottle #: 001

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	05:48:30
Calibration Chk	0.098	05:48:52
Air Blank	0.000	05:49:32

No RFI Detected