

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 03/25/2023

Test Time: 03:50

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 002

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:50:39
Calibration Chk	0.091	03:51:00
Air Blank	0.000	03:51:40

No RFI Detected

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 03/25/2023

Test Time: 03:44

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 002

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:44:51
Calibration Chk	0.083*	03:45:12
Air Blank	0.000	03:45:52

*Calibration Check Out of Tolerance

No RFI Detected

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 03/25/2023

Test Time: 03:42

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 002

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:43:03
Calibration Chk	0.085*	03:43:25
Air Blank	0.000	03:44:04

*Calibration Check Out of Tolerance

No RFI Detected

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY Goodyear P.D.
DATE 3/31/2023 TIME 0318
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.003 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample - Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS: QAS Completed

SIGNATURE K Miyazawa 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950331230231

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 03/31/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 03/31/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 03:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 002
Last Changed By: MIYAZATO, KORY

Permit #: 30215

RESULTS

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	03:18:32
Diagnostic Test	PASS	03:19:08
Air Blank	0.000	03:19:45
Calibration Chk	0.099	03:20:06
Air Blank	0.000	03:20:46
Subject Test 1	NSG*	03:23:59
Air Blank	0.000	03:24:41
Air Blank	0.000	03:25:21
Subject Test 2	0.000	03:25:44
Air Blank	0.000	03:26:25
Wait		03:30:29
Air Blank	0.000	03:31:06
Subject Test 3	INV**	03:31:27
Air Blank	RFI***	03:31:51

*No Sample Given

**Invalid Sample - Begin new deprivation period

***RFI Detect

EXCEPTION MESSAGES

* No Sample Given

** Invalid Sample - Begin new deprivation period

*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.

Consecutive subject test results must not differ by more than 0.020 g/210L.

Standard check results must be ±10%.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 03/31/2023

Test Time: 03:32

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 002

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:32:53
Calibration Chk	0.098	03:33:15
Air Blank	0.000	03:33:54

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950331230230

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 02/28/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 03/31/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 01:55
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 002
Last Changed By: MIYAZATO, KORY

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	03:14:30
Diagnostic Test	DIA*	03:14:36
Air Blank	0.000	03:15:13

*00004000000
IRPCM Status Error

Permit #: 30215

RESULTS

EXCEPTION MESSAGES

* Diagnostic Fail

OPERATOR COMMENTS

No RFI Detected

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 03/31/2023

Test Time: 03:16

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 002

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:16:36
Calibration Chk	0.099	03:16:57
Air Blank	0.000	03:17:39

No RFI Detected