

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY GOODYEAR
DATE 4/29/2023 TIME 0513
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration -.100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. 099 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazato 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950429230278

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 03/31/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 04/29/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 002
Last Changed By: MIYAZATO, KORY

Permit #: 30215

RESULTS

Test	g/210L	Time
Air Blank	0.000	05:13:05
Diagnostic Test	PASS	05:13:41
Air Blank	0.000	05:14:18
Calibration Chk	0.099	05:14:39
Air Blank	0.000	05:15:19
Subject Test 1	NSG*	05:21:22
Air Blank	0.000	05:22:06
Air Blank	0.000	05:22:47
Subject Test 2	0.000	05:23:06
Air Blank	0.000	05:23:47
Wait		05:27:51
Air Blank	0.000	05:28:29
Subject Test 3	INV**	05:28:47
Air Blank	RFI***	05:29:06

*No Sample Given
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* No Sample Given
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 04/29/2023

Test Time: 05:29

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 002

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	05:30:11
Calibration Chk	0.099	05:30:33
Air Blank	0.000	05:31:13

No RFI Detected