

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato AGENCY Goodyear
DATE 6/30/23 TIME 0406
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration -100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 099 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazato 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950630230372

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 06/30/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/30/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 03:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 003
Last Changed By: MIYAZATO, KORY

Permit #: 30215

RESULTS

Test	g/210L	Time
Air Blank	0.000	04:06:30
Diagnostic Test	PASS	04:07:06
Air Blank	0.000	04:07:43
Calibration Chk	0.099	04:08:05
Air Blank	0.000	04:08:44
Subject Test 1	NSG*	04:11:56
Air Blank	0.000	04:12:47
Air Blank	0.000	04:13:28
Subject Test 2	0.000	04:13:46
Air Blank	0.000	04:14:28
Wait		04:18:31
Air Blank	0.000	04:19:08
Subject Test 3	INV**	04:19:32
Air Blank	RFI***	04:19:46

*No Sample Given

**Invalid Sample - Begin new deprivation period

***RFI Detect

EXCEPTION MESSAGES

* No Sample Given

** Invalid Sample - Begin new deprivation period

*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.

Consecutive subject test results must not differ by more than 0.020 g/210L.

Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 06/30/2023

Test Time: 04:20

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 003

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:20:47
Calibration Chk	0.099	04:21:09
Air Blank	0.000	04:21:48

No RFI Detected

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST KM 42270 1139 AGENCY Gunnear
DATE 6/30/23 TIME 0339
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration 1.00 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0.00 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0.098 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ___ NO ___

COMMENTS:

MOUTH ALCOHOL TEST REGISTERED
- RANGE OCCURS.
- STARTED NOW I2

SIGNATURE [Signature]

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 06/30/2023

Test Time: 03:39

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 003

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	03:40:16
Calibration Chk	0.098	03:40:37
Air Blank	0.000	03:41:17

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950630230370

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 05/30/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/30/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 03:08
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 003
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	03:25:33
Diagnostic Test	PASS	03:26:09
Air Blank	0.000	03:26:46
Calibration Chk	0.099	03:27:07
Air Blank	0.000	03:27:47
Subject Test 1	NSG*	03:31:07
Air Blank	0.000	03:31:52
Air Blank	0.000	03:32:32
Subject Test 2	0.000	03:32:49
Air Blank	0.000	03:33:32
Wait		03:37:36
Air Blank	0.000	03:38:13
Subject Test 3	RNG**	03:38:29
Air Blank	RFI***	03:39:17

RESULTS

EXCEPTION MESSAGES

* No Sample Given
** Range Exceeded
*** RFI Detect

OPERATOR COMMENTS

*No Sample Given
**Range Exceeded
***RFI Detect

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950630230371

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 06/30/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/30/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 03:20
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 003
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	03:42:43
Diagnostic Test	PASS	03:43:19
Air Blank	0.000	03:43:56
Calibration Chk	0.099	03:44:17
Air Blank	0.000	03:44:57
Subject Test 1	NSG*	03:48:15
Air Blank	0.000	03:49:01
Air Blank	0.000	03:49:41
Subject Test 2	0.000	03:50:07
Air Blank	0.000	03:50:59
Wait		03:55:05
Air Blank	0.000	03:55:41
Subject Test 3	INV**	03:56:03
Air Blank	0.000	03:56:44

RESULTS

EXCEPTION MESSAGES

*No Sample Given
**Invalid Sample - Begin new deprivation period

* No Sample Given
** Invalid Sample - Begin new deprivation period

OPERATOR COMMENTS

No RFI Detected

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.