

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazaki AGENCY Goodyear
DATE 8/2/23 TIME 0216
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0.097 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Kurt 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950802230412

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 08/02/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 08/02/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 01:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 017
Last Changed By: MIYAZATO, KORY

Permit #: 30215

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	02:16:20
Diagnostic Test	PASS	02:16:56
Air Blank	0.000	02:17:33
Calibration Chk	0.098	02:17:54
Air Blank	0.000	02:18:34
Subject Test 1	NSG*	02:21:50
Air Blank	0.000	02:22:31
Air Blank	0.000	02:23:12
Subject Test 2	0.000	02:23:30
Air Blank	0.000	02:24:11
Wait		02:28:14
Air Blank	0.000	02:28:51
Subject Test 3	INV**	02:29:16
Air Blank	RFI***	02:29:32

RESULTS

EXCEPTION MESSAGES

*No Sample Given
**Invalid Sample - Begin new deprivation period
***RFI Detect

* No Sample Given
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 08/02/2023

Test Time: 02:30

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 017

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	02:30:24
Calibration Chk	0.097	02:30:45
Air Blank	0.000	02:31:25

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950802230411

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 06/30/2023
Last Annual Maintenance: 08/03/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 08/02/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 01:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	02:08:17
Diagnostic Test	DIA*	02:08:23
Air Blank	0.000	02:09:00

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 017
Last Changed By: MIYAZATO, KORY

*00004000000
IRPCM Status Error

Permit #: 30215

RESULTS

EXCEPTION MESSAGES

* Diagnostic Fail

OPERATOR COMMENTS

No RFI Detected

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 08/15/2023

Test Time: 23:50

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 017

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	23:50:45
Calibration Chk	0.099	23:51:07
Air Blank	0.000	23:51:47

No RFI Detected

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear P.D
DATE 8/31/2023 TIME 0513
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.00 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE  1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950831230435

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 08/02/2023
Last Annual Maintenance: 08/07/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 08/31/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:20
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 18022100A2
Expiration Date: 09/05/2024
Bottle #: 017
Last Changed By: MIYAZATO, KORY

Permit #: 30215

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	05:13:43
Diagnostic Test	PASS	05:14:19
Air Blank	0.000	05:14:58
Calibration Chk	0.099	05:15:19
Air Blank	0.000	05:15:59
Subject Test 1	DEF*	05:19:11
Air Blank	0.000	05:19:52
Air Blank	0.000	05:20:32
Subject Test 2	0.000	05:22:03
Air Blank	0.000	05:22:44
Wait		05:26:48
Air Blank	0.000	05:27:25
Subject Test 3	INV**	05:27:45
Air Blank	RFI***	05:28:04

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 08/31/2023

Test Time: 05:28

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 017

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	05:29:23
Calibration Chk	0.098	05:29:44
Air Blank	0.000	05:30:24

No RFI Detected