

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear PD
DATE 9/11/2023 TIME 2246
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazato 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050901230003

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 09/01/2023
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 09/01/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402690546
Expiration Date: 05/24/2025
Bottle #: 21
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	22:46:44
Diagnostic Test	PASS	22:47:20
Air Blank	0.000	22:47:57
Calibration Chk	0.101	22:48:19
Air Blank	0.000	22:48:58
Subject Test 1	DEF*	22:52:10
Air Blank	0.000	22:52:51
Air Blank	0.000	22:53:31
Subject Test 2	0.000	22:53:50
Air Blank	0.000	22:54:44
Wait		22:58:48
Air Blank	0.000	22:59:26
Subject Test 3	INV**	22:59:45
Air Blank	RFI***	22:59:58

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 09/01/2023

Test Time: 23:01

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402690546

Standard Value: 0.100

Expiration Date: 05/24/2025

Bottle #: 21

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	23:01:57
Calibration Chk	0.100	23:02:19
Air Blank	0.000	23:02:59

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050901230002

QAS:
QAS Permit #:
Agency:
Last 31-Day Check:
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 09/01/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402690546
Expiration Date: 05/24/2025
Bottle #: 21
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	21:44:33
Diagnostic Test	PASS	21:45:09
Air Blank	0.000	21:45:46
Calibration Chk	0.101	21:46:08
Air Blank	0.000	21:46:48
Subject Test 1	NSG*	21:50:00
Air Blank	0.000	21:52:07
Air Blank	0.000	21:52:47
Subject Test 2	0.000	21:53:32
Air Blank	0.000	21:54:15
Wait		21:58:19
Air Blank	0.000	21:58:56
Subject Test 3	INV**	21:59:42
Air Blank	RFI***	22:00:17

RESULTS

EXCEPTION MESSAGES

*No Sample Given
**Invalid Sample - Begin new deprivation period
***RFI Detect

* No Sample Given
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN 90-004605

Version - 9439.01.00

Test Date -9/1/2023 Test Time -21:32

Location -

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:32:46
Diagnostic Test	PASS	21:33:21

Extended Diag Info

Analytical Checksum PASS
Voltage/Current PASS
Memory PASS
Real Time Clock PASS
Temp Regulation PASS
ADC PASS
Analytical Status PASS
Version Status PASS
ITP PASS

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 09/01/2023

Test Time: 21:40

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402690546

Standard Value: 0.100

Expiration Date: 05/24/2025

Bottle #: 21

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:41:06
Calibration Chk	0.100	21:41:28
Air Blank	0.000	21:42:08

No RFI Detected

Successfully Completed Test Sequence

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 09/01/2023

Test Time: 21:30

Location:

Software Version: 9439.01.00

Standard Lot #:

Standard Value: 0.100

Expiration Date: 01/01/1900

Bottle #:

Last Changed By:

Permit #:

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:30:29
Calibration Chk	0.101	21:30:50
Air Blank	0.000	21:31:30

No RFI Detected

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

CMI
316 EAST NINTH STREET
OWENSBORO, KY 42303

INVOICE # : 1120627410
PO # : 8004678
CMI ITEM # : 340287
Date: May 24, 2023

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
PRODUCT EXPIRATION: May 24, 2025

CALGAZ LOT#: 302-402690546
ETHANOL IN NITROGEN

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

*Certification traceability is recognized by NIST (USA) through the CIPM (International Committee of Weights and Measures) MRA (Mutual Recognition Arrangement)

CALGAZ CYLINDER SIZE: 6DM

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST KM YAZATO 1139 AGENCY GOODYEAR PD
DATE 9/30/23 TIME 2122
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check


OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE  1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050930230016

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 09/01/2023
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 09/30/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402690546
Expiration Date: 05/24/2025
Bottle #: 21
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	21:22:49
Diagnostic Test	PASS	21:23:25
Air Blank	0.000	21:24:02
Calibration Chk	0.100	21:24:23
Air Blank	0.000	21:25:03
Subject Test 1	DEF*	21:28:14
Air Blank	0.000	21:28:55
Air Blank	0.000	21:29:35
Subject Test 2	0.000	21:29:54
Air Blank	0.000	21:30:34
Wait		21:34:39
Air Blank	0.000	21:35:16
Subject Test 3	INV**	21:35:35
Air Blank	RFI***	21:35:46

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 09/30/2023

Test Time: 21:36

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402690546

Standard Value: 0.100

Expiration Date: 05/24/2025

Bottle #: 21

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:36:38
Calibration Chk	0.100	21:37:00
Air Blank	0.000	21:37:39

No RFI Detected