



ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER MAINTENANCE REPORT

CR NO. I900037950005

Page 1 of 1

**AGENCY** Goodyear Police Department  
Goodyear, AZ 85338  
**OFFICER** MIYAZATO, #1139  
**DATE** October 18, 2023



**INSTRUMENT**

90003795

**INSTRUMENT LOCATION**

Goodyear Police Department

**MAINTENANCE REQUESTED**

Repair

**MAINTENANCE PERFORMED**

***Parts Replaced:***

Breath Hose

A handwritten signature in black ink, appearing to read "Kyle Jager".

---

KYLE JAGER, #7518, Forensic Scientist  
Central Regional Crime Laboratory  
2102 W. Encanto Blvd., Phoenix, AZ 85009  
602-223-2737 kjager@azdps.gov

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazaki AGENCY Goodyear PD  
DATE 10/31/23 TIME 2038  
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check  
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
5. Standard Calibration Check:  
0. 098 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGNATURE Kent 1139

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37951031230509

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 09/30/2023  
Last Annual Maintenance: 08/07/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 10/31/2023  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 20:00  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 18022100A2  
Expiration Date: 09/05/2024  
Bottle #: 015  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	20:38:31
Diagnostic Test	PASS	20:39:07
Air Blank	0.000	20:39:44
Calibration Chk	0.099	20:40:05
Air Blank	0.000	20:40:45
Subject Test 1	DEF*	20:43:56
Air Blank	0.000	20:44:44
Air Blank	0.000	20:45:24
Subject Test 2	0.000	20:45:41
Air Blank	0.000	20:46:23
Wait		20:50:27
Air Blank	0.000	20:51:04
Subject Test 3	INV**	20:51:30
Air Blank	RFI***	20:51:43

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 10/31/2023

Test Time: 20:52

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 015

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	20:52:47
Calibration Chk	0.098	20:53:08
Air Blank	0.000	20:53:48

No RFI Detected