

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato AGENCY GOODYEAR PD
DATE 10/31/23 TIME 2:23
INTOXILYZER SERIAL # 90-04605

1. Ensure that gas tank is attached and contains a standard alcohol concentration -100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazato 1139

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 10/31/2023

Test Time: 20:36

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402690546

Standard Value: 0.100

Expiration Date: 05/24/2025

Bottle #: 21

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	20:37:22
Calibration Chk	0.100	20:37:43
Air Blank	0.000	20:38:23

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46051031230047

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 09/30/2023
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 10/31/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 20:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402690546
Expiration Date: 05/24/2025
Bottle #: 21
Last Changed By: MIYAZATO, KORY

Permit #: 30215

RESULTS

Test	g/210L	Time
Air Blank	0.000	20:23:27
Diagnostic Test	PASS	20:24:04
Air Blank	0.000	20:24:41
Calibration Chk	0.100	20:25:02
Air Blank	0.000	20:25:41
Subject Test 1	DEF*	20:28:53
Air Blank	0.000	20:29:34
Air Blank	0.000	20:30:15
Subject Test 2	0.000	20:30:33
Air Blank	0.000	20:31:13
Wait		20:35:19
Air Blank	0.000	20:35:56
Subject Test 3	INV**	20:36:16
Air Blank	RFI***	20:36:30

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.