

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K. MIYAZATO AGENCY GOODYEAR P.D.  
DATE 11/30/23 TIME 0443  
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE K. Miyazato 1139

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37951130230539

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 10/31/2023  
Last Annual Maintenance: 08/07/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 11/30/2023  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 04:20  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 18022100A2  
Expiration Date: 09/05/2024  
Bottle #: 015  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	04:43:15
Diagnostic Test	PASS	04:43:50
Air Blank	0.000	04:44:27
Calibration Chk	0.099	04:44:49
Air Blank	0.000	04:45:28
Subject Test 1	DEF*	04:48:40
Air Blank	0.000	04:49:25
Air Blank	0.000	04:50:05
Subject Test 2	0.000	04:50:24
Air Blank	0.000	04:51:06
Wait		04:55:09
Air Blank	0.000	04:55:46
Subject Test 3	INV**	04:56:22
Air Blank	RFI***	04:56:38

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 11/30/2023

Test Time: 04:57

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 18022100A2

Standard Value: 0.100

Expiration Date: 09/05/2024

Bottle #: 015

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:57:30
Calibration Chk	0.098	04:57:52
Air Blank	0.000	04:58:31

No RFI Detected