

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K MIYAZATO AGENCY GOODYEAR P.D.
DATE 11/30/23 TIME 0447
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. 100 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE KM 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46051130230053

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 10/31/2023
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 11/30/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402690546
Expiration Date: 05/24/2025
Bottle #: 21
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	04:47:55
Diagnostic Test	PASS	04:48:31
Air Blank	0.000	04:49:08
Calibration Chk	0.100	04:49:29
Air Blank	0.000	04:50:08
Subject Test 1	DEF*	04:53:19
Air Blank	0.000	04:54:01
Air Blank	0.000	04:54:41
Subject Test 2	0.000	04:54:58
Air Blank	0.000	04:55:39
Wait		04:59:43
Air Blank	0.000	05:00:20
Subject Test 3	INV**	05:00:46
Air Blank	RFI***	05:00:59

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 11/30/2023

Test Time: 05:01

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402690546

Standard Value: 0.100

Expiration Date: 05/24/2025

Bottle #: 21

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	05:01:50
Calibration Chk	0.100	05:02:12
Air Blank	0.000	05:02:51

No RFI Detected