

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa AGENCY Goodyear
DATE 12/31/2003 TIME 0038
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration -100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazawa 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37951231230589

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 11/30/2023
Last Annual Maintenance: 08/07/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 12/31/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 00:20
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 011
Last Changed By: MIYAZATO, KORY

Permit #: 30215

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	00:38:22
Diagnostic Test	PASS	00:38:58
Air Blank	0.000	00:39:35
Calibration Chk	0.099	00:39:58
Air Blank	0.000	00:40:38
Subject Test 1	DEF*	00:43:49
Air Blank	0.000	00:44:30
Air Blank	0.000	00:45:10
Subject Test 2	0.000	00:45:28
Air Blank	0.000	00:46:09
Wait		00:50:14
Air Blank	0.000	00:50:51
Subject Test 3	INV**	00:51:38
Air Blank	RFI***	00:51:56

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 12/31/2023

Test Time: 00:52

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 011

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	00:52:48
Calibration Chk	0.100	00:53:10
Air Blank	0.000	00:53:49

No RFI Detected

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003795

Test Date: 12/29/2023

Test Time: 04:00

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 011

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank •	0.000	04:01:03
Calibration Chk	0.099	04:01:24
Air Blank	0.000	04:02:04

No RFI Detected

Successfully Completed Test Sequence



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 15958
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 24923100A1
Expiration: 10/5/2025

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	260 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 261.0 µmol/mol
Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

09-26-2023
Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.