

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050102240066
OAS: MIYAZATO, KORY
OAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 11/30/2023
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Test Date: 01/02/2024
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 11:05
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: GUILBAULT, MICHAEL
Agency: GOODYEAR PD
Permit #: 30305

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #:
Last Changed By:
GUILBAULT, MICHAEL
Permit #: 30305

Test g/210L Time
Air Blank 0.000 11:22:06
Diagnostic Test PASS 11:22:41
Air Blank 0.000 11:23:18
Calibration Chk 0.101 11:23:40
Air Blank 0.000 11:24:19
Subject Test 1 DEF* 11:27:35
Air Blank 0.000 11:28:46
Air Blank 0.000 11:29:27
Subject Test 2 0.000 11:29:49
Air Blank 0.000 11:30:32
Wait 11:34:35
Air Blank 0.000 11:35:12
Subject Test 3 INV** 11:35:40
Air Blank RFI*** 11:35:54

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 01/02/2024

Test Time: 11:37

Location: -----

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #:

Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Operator: GUILBAULT, MICHAEL

Permit #: 30305

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	11:37:36
Calibration Chk	0.101	11:37:58
Air Blank	0.000	11:38:37

No RFI Detected

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 01/02/2024

Test Time: 11:12

Location: -----

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #:

Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Operator: GUILBAULT, MICHAEL

Permit #: 30305

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	11:12:50
Calibration Chk	0.103	11:13:12
Air Blank	0.000	11:13:51

No RFI Detected

Successfully Completed Test Sequence

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Michael Guibault AGENCY GOODYEAR POLICE DEPT
DATE 01/02/2024 TIME 1122
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample -- Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. 101 AC
6. Air Blanks Completed
 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

New Standard charge - see Attached

SIGNATURE MG 1/4



7 Eastgate Dr. • P.O. Box 790 • Jacksonvillle, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 15958
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 24923100A1
Expiration: 10/5/2025

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
------------	-------------------------	-------------------------------	--------------------

Ethanol	260 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:

Certified Reference Material - 261.0 µmol/mol

Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).


Specialty Gas Lab Tech

09-26-2023
Issuance Date





The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory

ARIZONA FORENSIC BREATH ALCOHOL
ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050131240077
OAS: GUILBAULT, MICHAEL
OAS Permit #: 30305
Agency: GOODYEAR PD
Last 31-Day Check: 01/02/2024
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Test Date: 01/31/2024
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 21:00
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD
Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #:
Last Changed By:
GUILBAULT, MICHAEL
Permit #: 30305

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 01/31/2024

Test Time: 21:49

Location: -----

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #:

Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:49:43
Calibration Chk	0.100	21:50:05
Air Blank	0.000	21:50:44

Test	g/210L	Time
Air Blank	0.000	21:35:37
Diagnostic Test	PASS	21:36:13
Air Blank	0.000	21:36:50
Calibration Chk	0.101	21:37:11
Air Blank	0.000	21:37:51
Subject Test 1	DEF*	21:41:02
Air Blank	0.000	21:41:42
Air Blank	0.000	21:42:23
Subject Test 2	0.000	21:42:43
Air Blank	0.000	21:43:25
Wait		21:47:29
Air Blank	0.000	21:48:06
Subject Test 3	INV**	21:48:26
Air Blank	RFI***	21:48:41

No RFI Detected

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST KM1742AD AGENCY Goodyear
DATE 1/31/24 TIME 2135
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. 100 AC
6. Air Blanks Completed
 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE KM1742AD 1139